

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB133987

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/18/2024	08:00
2	CAL2	1	Water	NA	NA	20.2	7.00	12/18/2024	08:01
3	CAL3	1	Water	NA	NA	20.2	10.02	12/18/2024	08:03
4	ICV	1	Water	NA	NA	20.3	7.00	12/18/2024	08:05
5	CCV1	1	Water	NA	NA	20.3	2.02	12/18/2024	08:33
6	P5279-02	1	Solid	20.02	20	20.4	8.90	12/18/2024	08:40
7	P5279-02DUP	1	Solid	20.03	20	20.5	8.91	12/18/2024	08:41
8	CCV2	1	Water	NA	NA	20.3	12.02	12/18/2024	08:45

WORKLIST(Hardcopy Internal Chain)

VB 133987

WorkList Name : corrosivity p5279

WorkList ID : 186418

Department : Wet-Chemistry

Date : 12-18-2024 07:42:23

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5279-02	ROCKAWAY-PARK	Solid	Corrosivity	Cool 4 deg C	TULL02	L51	12/13/2024	9045D

Date/Time 12-18-24 07:50

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 12-18-24

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]