

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB133988

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3157

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	81.7	12/18/2024	08:40
2	P5313-01		1	11	100.00	>212.0	>212.0	12/18/2024	09:10
3	P5313-01DUP		1	11	100.00	>212.0	>212.0	12/18/2024	09:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB133988

WorkList Name : fp-12-17-

WorkList ID : 186401

Department : Wet-Chemistry

Date : 12-17-2024 13:16:48

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5313-01	FMI109	Water	Flash Point	Cool 4 deg C	PSEG03	L61	12/17/2024	1010B

Date/Time 12/18/2024 08:30
Raw Sample Received by: PM awc
Raw Sample Relinquished by: M wdc

Date/Time 12/18/2024 10:20
Raw Sample Received by: JGD
Raw Sample Relinquished by: RM wdc