

LB133989

7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

40450

A= 0.0000e+000

B= 2.4471e-004

C= -8.2640e-002

Rho= 0.9992420

Accept=Accepted

Max. Conc: — 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.014	-0.014	282	0.000	282	0				5
0.2	0.200	0.182	-0.018	1082	0.0 %	1082					1
2.5	2.500	2.526	0.026	10660	0.0 %	10660					2
5.0	5.000	4.907	-0.093	20388	0.0 %	20388					4
7.5	7.500	7.783	0.283	32142	0.0 %	32142					2
10.0	10.000	9.816	-0.184	40450	0.0 %	40450					

LB133989 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 18 Dec 2024 09:59:06

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	282	-	0.0000	7470A	PPB	18 Dec 2024 10:07:56	S	Std
0.2 - 1	50.2	1082	-	0.2000	7470A	PPB	18 Dec 2024 10:10:12	S	Std
2.5 - 1	52.5	10660	-	2.5000	7470A	PPB	18 Dec 2024 10:12:29	S	Std
5.0 - 1	55	20388	-	5.0000	7470A	PPB	18 Dec 2024 10:14:46	S	Std
7.5 - 1	57.5	32142	-	7.5000	7470A	PPB	18 Dec 2024 10:17:02	S	Std
10.0 - 1	510	40450	-	10.0000	7470A	PPB	18 Dec 2024 10:24:36	S	Std
ICV23 - 1	ICV23	15966	3.8244	-	7470A	PPB	18 Dec 2024 10:27:24	U	SMPL
ICB23 - 1	ICB23	161	-0.0432	-	7470A	PPB	18 Dec 2024 10:29:38	U	SMPL
CCV39 - 1	CCV39	20848	5.0191	-	7470A	PPB	18 Dec 2024 10:31:54	U	SMPL
CCB39 - 1	CCB39	127	-0.0516	-	7470A	PPB	18 Dec 2024 10:34:09	U	SMPL
CRA - 1	CRA	1065	0.1780	-	7470A	PPB	18 Dec 2024 10:36:26	U	SMPL
HighStd - 1	HighStd	40058	9.7201	-	7470A	PPB	18 Dec 2024 10:38:41	U	SMPL
ChkStd - 1	ChkStd	30422	7.3620	-	7470A	PPB	18 Dec 2024 10:40:55	U	SMPL
PB165718BL - 1	PBW	54	-0.0694	-	7470A	PPB	18 Dec 2024 10:43:13	U	SMPL
PB165718BS - 1	LCSW	14912	3.5665	-	7470A	PPB	18 Dec 2024 10:45:31	U	SMPL
P5313-01 - 1	FMI109	513	0.0429	-	7470A	PPB	18 Dec 2024 10:47:48	U	SMPL
P5313-01DUP - 1	FMI109DUP	549	0.0517	-	7470A	PPB	18 Dec 2024 10:50:05	U	SMPL
P5313-01MS - 1	FMI109MS	16120	3.8621	-	7470A	PPB	18 Dec 2024 10:52:22	U	SMPL
P5313-01MSD - 1	FMI109MSD	13730	3.2773	-	7470A	PPB	18 Dec 2024 10:54:39	U	SMPL
CCV40 - 1	CCV40	22080	5.3206	-	7470A	PPB	18 Dec 2024 10:56:55	U	SMPL
CCB40 - 1	CCB40	143	-0.0476	-	7470A	PPB	18 Dec 2024 10:59:12	U	SMPL
P5313-01LX5 - 1		157	-0.0442	-	7470A	PPB	18 Dec 2024 11:01:29	U	SMPL
P5313-01A - 1		15564	3.7261	-	7470A	PPB	18 Dec 2024 11:03:45	U	SMPL
CCV41 - 1	CCV41	19742	4.7485	-	7470A	PPB	18 Dec 2024 11:08:34	U	SMPL
CCB41 - 1	CCB41	175	-0.0398	-	7470A	PPB	18 Dec 2024 11:10:51	U	SMPL