

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB134005

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/19/2024	08:10
2	CAL2	1	Water	NA	NA	20.3	7.00	12/19/2024	08:11
3	CAL3	1	Water	NA	NA	20.3	10.02	12/19/2024	08:12
4	ICV	1	Water	NA	NA	20.3	7.00	12/19/2024	08:15
5	CCV1	1	Water	NA	NA	20.2	2.02	12/19/2024	08:20
6	P5341-01	1	Water	NA	NA	21.3	6.96	12/19/2024	08:25
7	P5341-03	1	Water	NA	NA	23.1	10.22	12/19/2024	08:35
8	P5343-06	1	Water	NA	NA	23.1	6.93	12/19/2024	08:44
9	P5343-06DUP	1	Water	NA	NA	23.2	6.94	12/19/2024	08:45
10	CCV2	1	Water	NA	NA	20.3	12.02	12/19/2024	08:50

U3 134005

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph p5341 WorkList ID : 186449 Department : Wet-Chemistry Date : 12-19-2024 07:44:28

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5341-01 <i>B</i>	STORMWATER-AQ-COMP	Water	pH	Cool 4 deg C	PSEG03	N12	12/18/2024	9040C
P5341-03 <i>B</i>	TOTE-425	Water	pH	Cool 4 deg C	PSEG03	N12	12/18/2024	9040C
P5343-06 <i>B</i>	PUMP-WATER-271	Water	pH	Cool 4 deg C	PSEG03	N12	12/18/2024	9040C

Date/Time 12-19-24 08:00

Raw Sample Received by: JB WOC

Raw Sample Relinquished by: clgr

Date/Time 12-19-24 10:30

Raw Sample Received by: clgr

Raw Sample Relinquished by: JB WOC