

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB134006

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/19/2024	08:10
2	CAL2	1	Water	NA	NA	20.3	7.00	12/19/2024	08:11
3	CAL3	1	Water	NA	NA	20.3	10.02	12/19/2024	08:12
4	ICV	1	Water	NA	NA	20.3	7.00	12/19/2024	08:15
5	CCV1	1	Water	NA	NA	20.1	2.01	12/19/2024	09:05
6	P5330-04	1	Solid	20.02	20	22.2	7.67	12/19/2024	09:15
7	P5330-04DUP	1	Solid	20.03	20	22.3	7.68	12/19/2024	09:16
8	P5341-02	1	Solid	20.02	20	23.4	7.09	12/19/2024	09:25
9	P5343-05	1	Solid	20.03	20	23.6	6.60	12/19/2024	09:30
10	CCV2	1	Water	NA	NA	20.3	12.02	12/19/2024	09:31

WORKLIST(Hardcopy Internal Chain)

W3134006

WorkList Name : corrosivity p5330

WorkList ID : 188450

Department : Wet-Chemistry

Date : 12-19-2024 07:44:59

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5330-04	TP-5	Solid	Corrosivity	Cool 4 deg C	PSEG03	L51	12/18/2024	9045D
P5341-02	STORMWATER-SOLID-COMP	Solid	Corrosivity	Cool 4 deg C	PSEG03	N12	12/18/2024	9045D
P5343-05	OILY-RAGS-274	Solid	Corrosivity	Cool 4 deg C	PSEG03	N12	12/18/2024	9045D

Date/Time 12-19-24 08:00
 Raw Sample Received by: H. Wells
 Raw Sample Relinquished by: C. Smith

Date/Time 12-19-24 11:00
 Raw Sample Received by: C. Smith
 Raw Sample Relinquished by: H. Wells