

LB134019


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

5712E



A= 0.0000e+000

B= 1.7897e-004

C= 1.1080e-002

Rho= 0.9989323

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Y.D
0.0	0.000	0.054	0.054	240	0.000	240					1
0.2	0.200	0.236	0.036	1256	0.0 %	1256					18
2.5	2.500	2.622	0.122	14587	0.0 %	14587					5
5.0	5.000	4.794	-0.206	26726	0.0 %	26726					-4
7.5	7.500	7.259	-0.241	40500	0.0 %	40500					-3
10.0	10.000	10.235	0.235	57125	0.0 %	57125					2

LB134019 INMSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 19 Dec 2024 15:13:26

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	240	-	0.0000	7471B	PPB	19 Dec 2024 15:17:09	S	Std
0.2 - 1	50-2	1256	-	0.2000	7471B	PPB	19 Dec 2024 15:19:26	S	Std
2.5 - 1	52-5	14587	-	2.5000	7471B	PPB	19 Dec 2024 15:21:43	S	Std
5.0 - 1	55	26726	-	5.0000	7471B	PPB	19 Dec 2024 15:24:00	S	Std
7.5 - 1	57-5	40500	-	7.5000	7471B	PPB	19 Dec 2024 15:29:31	S	Std
10.0 - 1	96	57125	-	10.0000	7471B	PPB	19 Dec 2024 15:31:46	S	Std
ICV26 - 1	ICV26	24149	4.3330	-7471B	PPB	19 Dec 2024 15:40:00	U	SMPL	
ICB26 - 1	ICB26	-11	0.0091	-7471B	PPB	19 Dec 2024 15:42:17	U	SMPL	
CCV51 - 1	CCV51	27452	4.9242	-7471B	PPB	19 Dec 2024 15:44:35	U	SMPL	
CCB51 - 1	CCB51	-64	-0.0004	-7471B	PPB	19 Dec 2024 15:46:50	U	SMPL	
CRA - 1	CRA	1215	0.2285	-7471B	PPB	19 Dec 2024 15:49:07	U	SMPL	
HighStd - 1	HighStd	58020	10.3949	-7471B	PPB	19 Dec 2024 15:54:04	U	SMPL	
ChkStd - 1	ChkStd	36843	6.6049	-7471B	PPB	19 Dec 2024 15:56:19	U	SMPL	
PB165774BL - 1	PBS	207	0.0481	-7471B	PPB	19 Dec 2024 15:58:41	U	SMPL	
PB165774BS - 1	LCSS	20704	3.7165	-7471B	PPB	19 Dec 2024 16:06:53	U	SMPL	
P5330-01 - 1	TP-5	803	0.1548	-7471B	PPB	19 Dec 2024 16:09:09	U	SMPL	
P5330-01DUP - 1	TP-5DUP	982	0.1868	-7471B	PPB	19 Dec 2024 16:11:29	U	SMPL	
P5330-01MS - 1	TP-5MS	19549	3.5098	-7471B	PPB	19 Dec 2024 16:16:46	U	SMPL	
P5330-01MSD - 1	TP-5MSD	19601	3.5191	-7471B	PPB	19 Dec 2024 16:19:02	U	SMPL	
P5342-01 - 1	CHRT26634	23673	4.2478	-7471B	PPB	19 Dec 2024 16:21:20	U	SMPL	
CCV52 - 1	CCV52	27396	4.9142	-7471B	PPB	19 Dec 2024 16:23:40	U	SMPL	
CCB52 - 1	CCB52	-72	-0.0018	-7471B	PPB	19 Dec 2024 16:26:07	U	SMPL	
P5342-03 - 1	HT2651	7562	1.3645	-7471B	PPB	19 Dec 2024 16:28:25	U	SMPL	
P5342-05 - 1	RB21198	1574	0.2928	-7471B	PPB	19 Dec 2024 16:30:40	U	SMPL	
P5343-01 - 1	VNJ210	3673	0.6684	-7471B	PPB	19 Dec 2024 16:33:00	U	SMPL	
P5343-03 - 1	VNJ281	7954	1.4346	-7471B	PPB	19 Dec 2024 16:35:16	U	SMPL	
P5330-01LX5 - 1		114	0.0315	-7471B	PPB	19 Dec 2024 16:37:33	U	SMPL	
P5330-01A - 1		19854	3.5644	-7471B	PPB	19 Dec 2024 16:39:52	U	SMPL	
CCV53 - 1	CCV53	26626	4.7763	-7471B	PPB	19 Dec 2024 16:42:08	U	SMPL	
CCB53 - 1	CCB53	-38	0.0043	-7471B	PPB	19 Dec 2024 16:44:28	U	SMPL	