

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB134027

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/20/2024	08:05
2	CAL2	1	Water	NA	NA	20.2	7.00	12/20/2024	08:06
3	CAL3	1	Water	NA	NA	20.3	10.02	12/20/2024	08:09
4	ICV	1	Water	NA	NA	20.3	7.01	12/20/2024	08:11
5	CCV1	1	Water	NA	NA	20.3	2.01	12/20/2024	08:15
6	P5356-01	1	Water	NA	NA	20.1	6.66	12/20/2024	08:25
7	P5356-01DUP	1	Water	NA	NA	20.2	6.67	12/20/2024	08:26
8	CCV2	1	Water	NA	NA	20.3	12.02	12/20/2024	08:30

WORKLIST(Hardcopy Internal Chain)

VB 134024

WorkList Name : ph p5346

WorkList ID : 186505

Department : Wet-Chemistry

Date : 12-20-2024 07:51:51

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5356-01	2 STAND-PIPE	Water	pH	Cool 4 deg C	PSEG03	N13	12/19/2024	9040C

Date/Time 12/20/24 08:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 12/20/24

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]