

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB134028

Slope : 98.4

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/20/2024	08:05
2	CAL2	1	Water	NA	NA	20.2	7.00	12/20/2024	08:06
3	CAL3	1	Water	NA	NA	20.3	10.02	12/20/2024	08:09
4	ICV	1	Water	NA	NA	20.3	7.01	12/20/2024	08:11
5	CCV1	1	Water	NA	NA	20.4	2.02	12/20/2024	08:35
6	P5355-02	1	Solid	20.02	20	24.3	6.75	12/20/2024	08:45
7	P5355-02DUP	1	Solid	20.03	20	24.4	6.76	12/20/2024	08:46
8	CCV2	1	Water	NA	NA	20.3	12.02	12/20/2024	08:50

WORKLIST(Hardcopy Internal Chain)

UB 134028

WorkList Name : corrosivity p5355

WorkList ID : 186508

Department : Wet-Chemistry

Date : 12-20-2024 07:53:54

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5355-02	RBR251688	Solid	Corrosivity	Cool 4 deg C	PSEG03	N13	12/19/2024	9045D

Date/Time 12/20/24 08:00

Raw Sample Received by: JF wey

Raw Sample Relinquished by: af sm

Date/Time 12/20/24 11:30

Raw Sample Received by: af sm

Raw Sample Relinquished by: JF wey