

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB134033

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3157

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	12/23/2024	09:00
2	P5356-01		1	14	100.00	>212.0	>212.0	12/23/2024	09:30
3	P5356-01DUP		1	14	100.00	>212.0	>212.0	12/23/2024	10:00
4	P5357-01		1	12	100.00	>212.0	>212.0	12/23/2024	10:30
5	P5357-03		1	14	100.00	>212.0	>212.0	12/23/2024	11:00
6	P5377-02		1	13	100.00	>212.0	>212.0	12/23/2024	12:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp-12-20-

WorkList ID : 186513

Department : Wet-Chemistry

Date : 12-20-2024 08:20:54

66134033

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5356-01	STAND-PIPE	Water	Flash Point	Cool 4 deg C	PSEG03	N13	12/19/2024	1010B
P5357-01	TOTE-2062	Water	Flash Point	Cool 4 deg C	PSEG03	N13	12/19/2024	1010B
P5357-03	121824-C	Water	Flash Point	Cool 4 deg C	PSEG03	N13	12/19/2024	1010B
P5377-02	GAS-TRE-1114	Water	Flash Point	Cool 4 deg C	PSEG03	N13	12/20/2024	1010B

Date/Time 12/23/2024 08:20
 Raw Sample Received by: RM CWG
 Raw Sample Relinquished by: RB (RM)

Date/Time 12/23/2024 12:46
 Raw Sample Received by: RB (RM)
 Raw Sample Relinquished by: RM CWG