

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB134055

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/21/2024	09:30
2	CAL2	1	Water	NA	NA	20.3	7.00	12/21/2024	09:31
3	CAL3	1	Water	NA	NA	20.3	10.02	12/21/2024	09:33
4	ICV	1	Water	NA	NA	20.3	7.01	12/21/2024	09:37
5	CCV1	1	Water	NA	NA	20.3	2.01	12/21/2024	09:40
6	P5318-01	1	Solid	20.02	20	22.6	8.27	12/21/2024	09:50
7	P5318-01DUP	1	Solid	20.03	20	22.7	8.28	12/21/2024	09:51
8	CCV2	1	Water	NA	NA	20.3	12.02	12/21/2024	09:55

WORKLIST(Hardcopy Internal Chain)

✓ 134055

WorkList Name : ph p5318 WorkList ID : 186548 Department : Wet-Chemistry Date : 12-21-2024 09:18:11

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5318-01	✓ AU-06-122024	Solid	pH	Cool 4 deg C	PSEG05	N31	12/20/2024	9045D

Date/Time 12-21-24 09:12:00
Raw Sample Received by: 3860001
Raw Sample Relinquished by: 3860001

12-21-24 11:00
Date/Time 12-21-24
Raw Sample Received by: 3860001
Raw Sample Relinquished by: 3860001