

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB134056

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/21/2024	09:20
2	CAL2	1	Water	NA	NA	20.3	7.00	12/21/2024	09:31
3	CAL3	1	Water	NA	NA	20.3	10.02	12/21/2024	09:33
4	ICV	1	Water	NA	NA	20.3	7.01	12/21/2024	09:37
5	CCV1	1	Water	NA	NA	20.3	2.01	12/21/2024	09:40
6	P5377-02	1	Water	NA	NA	20.6	7.50	12/21/2024	10:10
7	P5377-02DUP	1	Water	NA	NA	20.7	7.51	12/21/2024	10:11
8	CCV2	1	Water	NA	NA	20.3	12.02	12/21/2024	10:15

WORKLIST(Hardcopy Internal Chain)

VP 134056

WorkList Name : ph water p5377

WorkList ID : 186549

Department : Wet-Chemistry

Date : 12-21-2024 09:19:03

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5377-02	GAS-TRE-1114	Water	pH	Cool 4 deg C	PSEG03	N13	12/20/2024	9040C

Date/Time 12.21.24 09:20
Raw Sample Received by: JB (WCC)
Raw Sample Relinquished by: JB (WCC)

Date/Time 12.21.24 11:00
Raw Sample Received by: JB (WCC)
Raw Sample Relinquished by: JB (WCC)