

LB134003

7470A

INSTRUMENT ID: (VJ)

Linear

 μ Abs.:

52266

A= 0.0000e+000

B= 1.9023e-004

C= -3.8559e-002

Rho= 0.9998584

Accept=Accepted

Max. Conc: — 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.020	-0.020	96	0.000	96	0			
0.2	0.200	0.171	-0.029	1100	0.0 %	1100				
2.5	2.500	2.523	0.023	13467	0.0 %	13467				
5.0	5.000	5.013	0.013	26556	0.0 %	26556				
7.5	7.500	7.609	0.109	40203	0.0 %	40203				
10.0	10.000	9.904	-0.096	52266	0.0 %	52266				

0/02
-15
-10
-1

LB134063 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 Dec 2024 11:59:25

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	96	-	0.0000	7470A	PPB	23 Dec 2024 12:06:12	S	Std
0.2 - 1	50.2	1100	-	0.2000	7470A	PPB	23 Dec 2024 12:10:30	S	Std
2.5 - 1	52.5	13467	-	2.5000	7470A	PPB	23 Dec 2024 12:12:47	S	Std
5.0 - 1	55	26556	-	5.0000	7470A	PPB	23 Dec 2024 12:15:04	S	Std
7.5 - 1	57.5	40203	-	7.5000	7470A	PPB	23 Dec 2024 12:17:21	S	Std
10.0 - 1	511	52266	-	10.0000	7470A	PPB	23 Dec 2024 12:25:02	S	Std
ICV29 - 1	ICV29	21557	4.0622	-	7470A	PPB	23 Dec 2024 12:33:32	U	SMPL
ICB29 - 1	ICB29	214	0.0022	-	7470A	PPB	23 Dec 2024 12:35:47	U	SMPL
CCV60 - 1	CCV60	28457	5.3748	-	7470A	PPB	23 Dec 2024 12:38:03	U	SMPL
CCB60 - 1	CCB60	145	-0.0110	-	7470A	PPB	23 Dec 2024 12:40:18	U	SMPL
CRA - 1	CRA	1188	0.1874	-	7470A	PPB	23 Dec 2024 12:42:35	U	SMPL
HighStd - 1	HighStd	52474	9.9435	-	7470A	PPB	23 Dec 2024 12:44:50	U	SMPL
ChkStd - 1	ChkStd	37188	7.0357	-	7470A	PPB	23 Dec 2024 12:47:04	U	SMPL
PB165802BL - 1	PBW	-319	-0.0992	-	7470A	PPB	23 Dec 2024 12:49:23	U	SMPL
PB165802BS - 1	LCSW	20160	3.7965	-	7470A	PPB	23 Dec 2024 12:51:41	U	SMPL
P5330-04 - 1	TP-5	-127	-0.0627	-	7470A	PPB	23 Dec 2024 12:53:58	U	SMPL
P5338-01 - 1	USED-OIL-DEBRIS	81	-0.0232	-	7470A	PPB	23 Dec 2024 12:56:16	U	SMPL
P5338-02 - 1	USED-MOBILE-TRANSFORMER-OIL	213	0.0020	-	7470A	PPB	23 Dec 2024 12:58:33	U	SMPL
P5341-02 - 1	STORMWATER-SOLID-COMP	-1	-0.0387	-	7470A	PPB	23 Dec 2024 13:00:49	U	SMPL
P5343-05 - 1	OILY-RAGS-274	1426	0.2327	-	7470A	PPB	23 Dec 2024 13:03:04	U	SMPL
CCV61 - 1	CCV61	26594	5.0204	-	7470A	PPB	23 Dec 2024 13:05:19	U	SMPL
CCB61 - 1	CCB61	158	-0.0085	-	7470A	PPB	23 Dec 2024 13:07:34	U	SMPL
P5355-02 - 1	RBR251688	678	0.0904	-	7470A	PPB	23 Dec 2024 13:09:51	U	SMPL
P5355-02DUP - 1	RBR251688DUP	498	0.0562	-	7470A	PPB	23 Dec 2024 13:12:06	U	SMPL
P5355-02MS - 1	RBR251688MS	19967	3.7597	-	7470A	PPB	23 Dec 2024 13:14:21	U	SMPL
P5355-02MSD - 1	RBR251688MSD	21348	4.0225	-	7470A	PPB	23 Dec 2024 13:16:37	U	SMPL
PB165804BS - 1	LCSW	19430	3.6576	-	7470A	PPB	23 Dec 2024 13:21:13	U	SMPL
P5341-01 - 1	STORMWATER-AQ-COMP	4266	0.7730	-	7470A	PPB	23 Dec 2024 13:23:29	U	SMPL
P5341-03 - 1	TOTE-425	1634	0.2723	-	7470A	PPB	23 Dec 2024 13:25:48	U	SMPL
P5343-06 - 1	PUMP-WATER-271	1332	0.2148	-	7470A	PPB	23 Dec 2024 13:28:08	U	SMPL
P5356-01 - 1	STAND-PIPE	1925	0.3278	-	7470A	PPB	23 Dec 2024 13:31:24	U	SMPL
CCV62 - 1	CCV62	28070	5.3012	-	7470A	PPB	23 Dec 2024 13:41:28	U	SMPL
CCB62 - 1	CCB62	8	-0.0370	-	7470A	PPB	23 Dec 2024 13:43:44	U	SMPL
P5377-02 - 1	GAS-TRE-1114	342	0.0265	-	7470A	PPB	23 Dec 2024 13:46:02	U	SMPL
P5377-02DUP - 1	GAS-TRE-1114DUP	425	0.0423	-	7470A	PPB	23 Dec 2024 13:48:18	U	SMPL
P5377-02MS - 1	GAS-TRE-1114MS	18196	3.4229	-	7470A	PPB	23 Dec 2024 14:00:28	U	SMPL
P5377-02MSD - 1	GAS-TRE-1114MSD	21006	3.9574	-	7470A	PPB	23 Dec 2024 14:02:43	U	SMPL
PB165733TB - 1	PB165733TB	-199	-0.0764	-	7470A	PPB	23 Dec 2024 14:05:00	U	SMPL
P5355-02LX5 - 1		24	-0.0340	-	7470A	PPB	23 Dec 2024 14:07:19	U	SMPL
P5355-02A - 1		18907	3.5581	-	7470A	PPB	23 Dec 2024 14:09:35	U	SMPL
P5377-02LX5 - 1		-322	-0.0998	-	7470A	PPB	23 Dec 2024 14:11:52	U	SMPL
P5377-02A - 1		19675	3.7042	-	7470A	PPB	23 Dec 2024 14:14:11	U	SMPL
PB165804BL - 1	PBW	-73	-0.0524	-	7470A	PPB	23 Dec 2024 14:16:27	U	SMPL
CCV63 - 1	CCV63	25365	4.7866	-	7470A	PPB	23 Dec 2024 14:18:44	U	SMPL
CCB63 - 1	CCB63	49	-0.0292	-	7470A	PPB	23 Dec 2024 14:21:00	U	SMPL