



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 12/27/2024

OVENTEMP IN Celsius(°C): 106
Time IN: 12:25
In Date: 12/26/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:25
Out Date: 12/27/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLIDS-OVEN

QC:LB134071

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P5322-01	YE5X5	1	1.15	8.43	9.58	7.81	79.0	
P5322-02	YE5X6	2	1.18	8.66	9.84	8.21	81.2	
P5322-03	YE5X6MS	3	1.18	8.66	9.84	8.21	81.2	
P5322-04	YE5X6MSD	4	1.18	8.66	9.84	8.21	81.2	
P5322-05	YE5X7	5	1.15	8.49	9.64	8.1	81.9	
P5322-06	YE5X8	6	1.16	8.64	9.8	8.01	79.3	
P5322-07	YE5X9	7	1.16	8.63	9.79	8.34	83.2	
P5322-08	YE5Y1	8	1.15	8.50	9.65	7.94	79.9	
P5322-09	YE5Z7	9	1.19	8.68	9.87	8.93	89.2	
P5322-10	YE5Z8	10	1.14	8.40	9.54	8.03	82.0	
P5322-11	YE5Z9	11	1.18	8.43	9.61	7.64	76.6	
P5322-12	YE600	12	1.15	8.80	9.95	8.36	81.9	
P5322-13	YE601	13	1.15	8.84	9.99	8.63	84.6	
P5322-14	YE695	14	1.16	8.80	9.96	8.51	83.5	
P5322-15	VHBLK001	15	1.00	1.00	2.00	2.00	100.0	VHBLK
P5322-16	YE630	16	1.13	8.45	9.58	8.41	86.2	
P5322-17	YE631	17	1.19	8.53	9.72	7.55	74.6	
P5322-18	YE635	18	1.14	8.62	9.76	8.18	81.7	
P5322-19	YE642	19	1.15	8.50	9.65	8.63	88.0	
P5322-20	YE643	20	1.14	8.85	9.99	8.48	82.9	
P5322-21	YE644	21	1.19	8.62	9.81	7.9	77.8	
P5322-22	YE645	22	1.15	8.38	9.53	8.07	82.6	
P5322-23	YE646	23	1.11	8.74	9.85	8.39	83.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$