

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB134091

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/26/2024	16:00
2	CAL2	1	Water	NA	NA	20.2	7.00	12/26/2024	16:01
3	CAL3	1	Water	NA	NA	20.2	10.02	12/26/2024	16:05
4	ICV	1	Water	NA	NA	20.3	7.01	12/26/2024	16:10
5	CCV1	1	Water	NA	NA	20.2	2.02	12/26/2024	16:11
6	P5386-05	1	Water	NA	NA	20.1	7.40	12/26/2024	16:20
7	P5386-05DUP	1	Water	NA	NA	20.2	7.41	12/26/2024	16:21
8	CCV2	1	Water	NA	NA	20.3	12.02	12/26/2024	16:25

WORKLIST(Hardcopy Internal Chain)

UB 134091

WorkList Name : ph water p5386

WorkList ID : 186632

Department : Wet-Chemistry

Date : 12-26-2024 15:45:03

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5386-05 ✓	MOO-24-00397	Water	pH	Cool 4 deg C	PSEG03	N31	12/26/2024	9040C

Date/Time

12/26/24 15:50

Raw Sample Received by:

JD WOC

Raw Sample Relinquished by:

JD WOC

Date/Time

12/26/24

Raw Sample Received by:

JD WOC

Raw Sample Relinquished by:

JD WOC