

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB134092

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/26/2024	16:00
2	CAL2	1	Water	NA	NA	20.2	7.00	12/26/2024	16:01
3	CAL3	1	Water	NA	NA	20.2	10.02	12/26/2024	16:05
4	ICV	1	Water	NA	NA	20.3	7.01	12/26/2024	16:10
5	CCV1	1	Water	NA	NA	20.3	2.01	12/26/2024	16:30
6	P5362-02	1	Solid	20.02	20	20.9	7.60	12/26/2024	16:37
7	P5362-02DUP	1	Solid	20.03	20	20.8	7.61	12/26/2024	16:38
8	P5386-02	1	Solid	20.02	20	21.9	6.72	12/26/2024	16:45
9	P5386-04	1	Solid	20.01	20	21.9	6.80	12/26/2024	16:50
10	CCV2	1	Water	NA	NA	20.3	12.02	12/26/2024	17:00

WORKLIST(Hardcopy Internal Chain)

VB 134092

WorkList Name : corrosivity p5387 WorkList ID : 186631 Department : Wet-Chemistry Date : 12-26-2024 15:44:36

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5362-02	WC-SOIL-20241219	Solid	Corrosivity	Cool 4 deg C	PARS02	N21	12/19/2024	9045D
P5386-02	MOO-24-00398	Solid	Corrosivity	Cool 4 deg C	PSEG03	N31	12/26/2024	9045D
P5386-04	MOO-24-00395-96	Solid	Corrosivity	Cool 4 deg C	PSEG03	N31	12/26/2024	9045D

Date/Time 12/26/24 15:50
Raw Sample Received by: JB WOC
Raw Sample Relinquished by: JB WOC

Date/Time 12/26/24 18:00
Raw Sample Received by: JB WOC
Raw Sample Relinquished by: JB WOC