

Analytical Summary Report

Analysis Method: SM3500-Cr B

ANALYST: rubina

Parameter: Hexavalent Chromium

SUPERVISOR REVIEW BY: Iwona

Run Number: LB134095

pH Meter ID: WC pH Meter-1

Reagent/Standard	Lot/Log #
Calibration Std. hexchrome 0.1 ppm	WP111231
Calibration Std. hexchrome 0.05 ppm	WP111230
calibration std. hexchrome 0.01 ppm	WP111228
calibration std. hexchrome 0 ppm	WP111227
hexavalent chromium color reagent	WP111205
0.2N SULFURIC ACID	WP109325
Calibration Std Hexachrome 0.025 ppm	WP111229
Hexavalent Chromium ICV-LCS Std	WP111234
Calibration and CCV std HexChrome 0.5PPM	WP111232
Calibration std HexChrome 1.0PPM	WP111233

Intercept: -0.0004

Slope: 0.7825

Regression: 0.999998

Seq	Lab ID	True Value (mg/l)	DF	Initial Vol (ml)	Final Vol (ml)	pH HN03	pH H2SO4	Absorb.at 540nm		Absorbance Difference	Result (mg/L)	%D	Anal Date	Anal Time
								Backgrnd	Color					
1	CAL1	0	1	100	100		1.77	0.000	0.000	0.000	0.000		12/26/2024	17:00
2	CAL2	0.01	1	100	100		1.91	0.000	0.007	0.007	0.009	-10	12/26/2024	17:01
3	CAL3	0.025	1	100	100		1.87	0.000	0.019	0.019	0.024	-4	12/26/2024	17:02
4	CAL4	0.05	1	100	100		1.85	0.000	0.038	0.038	0.049	-2	12/26/2024	17:03
5	CAL5	0.1	1	100	100		1.87	0.000	0.079	0.079	0.101	1	12/26/2024	17:04
6	CAL6	0.5	1	100	100		1.89	0.000	0.391	0.391	0.500	0	12/26/2024	17:05
7	CAL7	1	1	100	100		1.90	0.000	0.782	0.782	0.999	-0.1	12/26/2024	17:06

Analytical Summary Report

Analysis Method: SM3500-Cr B

ANALYST: rubina

Parameter: Hexavalent Chromium

SUPERVISOR REVIEW BY: Iwona

Run Number: LB134095

pH Meter ID: WC pH Meter-1

Seq	Lab ID	True Value	DF	Initial Vol (ml/gm)	Final Vol (ml)	pH HN03	pH H2SO4	Absorb.at540nm		Absorbance Difference	Intermediate Result (mg/L)	Anal Date	Anal Time
								Backgrnd	Color				
1	ICV	0.5	1	100	100		1.95	0.000	0.392	0.392	0.501	12/26/2024	17:07
2	ICB		1	100	100		1.79	0.000	0.000	0.000	0.001	12/26/2024	17:08
3	CCV1	0.5	1	100	100		1.96	0.000	0.390	0.390	0.499	12/26/2024	17:09
4	CCB1		1	100	100		1.76	0.000	0.000	0.000	0.001	12/26/2024	17:10
5	RL Check	0.01	1	100	100		1.90	0.000	0.008	0.008	0.011	12/26/2024	17:11
6	LB134095BL		1	100	100		1.77	0.000	0.000	0.000	0.001	12/26/2024	17:12
7	LB134095BS	0.5	1	100	100		1.92	0.000	0.399	0.399	0.510	12/26/2024	17:13
8	P5386-05		1	100	100		2.08	0.000	0.000	0.000	0.001	12/26/2024	17:14
9	P5386-05DU		1	100	100		2.10	0.000	0.000	0.000	0.001	12/26/2024	17:15
10	P5386-05MS	1	2	100	100		2.10	0.000	0.377	0.377	0.482	12/26/2024	17:16
11	P5386-05MS	1	2	100	100		2.13	0.000	0.379	0.379	0.485	12/26/2024	17:17
12	CCV2	0.5	1	100	100		1.94	0.000	0.392	0.392	0.501	12/26/2024	17:18
13	CCB2		1	100	100		1.75	0.000	0.000	0.000	0.001	12/26/2024	17:19

WORKLIST(Hardcopy Internal Chain)

66134095

WorkList Name : HEX-12-26.

WorkList ID : 186626

Department : Wet-Chemistry

Date : 12-26-2024 13:13:33

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5386-05	MOO-24-00397	Water	Hexavalent Chromium	Ammonium sulfate buffer	PSEG03	N31	12/26/2024	SM3500-Cr B

Date/Time 12/26/2024 15:00
Raw Sample Received by: RM CWJ
Raw Sample Relinquished by: RM CWJ

Date/Time 12/26/2024 16:30
Raw Sample Received by: RM CWJ
Raw Sample Relinquished by: RM CWJ