

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB134110

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/27/2024	13:44
2	CAL2	1	Water	NA	NA	20.3	7.00	12/27/2024	13:45
3	CAL3	1	Water	NA	NA	20.3	10.02	12/27/2024	13:50
4	ICV	1	Water	NA	NA	20.1	7.00	12/27/2024	14:00
5	CCV1	1	Water	NA	NA	20.3	2.01	12/27/2024	14:03
6	P5362-01	1	Solid	20.02	20	20.3	7.60	12/27/2024	14:10
7	P5362-01DUP	1	Solid	20.03	20	20.4	7.61	12/27/2024	14:11
8	CCV2	1	Water	NA	NA	20.2	12.02	12/27/2024	14:15

WORKLIST(Hardcopy Internal Chain)

134110

WorkList Name : ph p5362 WorkList ID : 186667 Department : Wet-Chemistry Date : 12-27-2024 13:32:12

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5362-01	WC-SOIL-20241219	Solid	pH	Cool 4 deg C	PARS02	N21	12/19/2024	9045D

Date/Time 12/27/24 13:40
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 12/24/24 16:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]