

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB134115

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/29/2024	08:30
2	CAL2	1	Water	NA	NA	20.2	7.00	12/29/2024	08:31
3	CAL3	1	Water	NA	NA	20.2	10.02	12/29/2024	08:32
4	ICV	1	Water	NA	NA	20.3	7.02	12/29/2024	08:37
5	CCV1	1	Water	NA	NA	20.3	2.01	12/29/2024	08:40
6	P5389-02	1	Water	NA	NA	20.4	7.50	12/29/2024	08:47
7	P5389-02DUP	1	Water	NA	NA	20.5	7.51	12/29/2024	08:48
8	P5389-03	1	Water	NA	NA	20.6	7.58	12/29/2024	08:55
9	P5389-05	1	Water	NA	NA	20.9	7.54	12/29/2024	09:00
10	P5389-06	1	Water	NA	NA	20.8	7.93	12/29/2024	09:05
11	CCV2	1	Water	NA	NA	20.3	12.02	12/29/2024	09:10

WORKLIST(Hardcopy Internal Chain)

134115

WorkList Name : ph water p5389

WorkList ID : 186668

Department : Wet-Chemistry

Date : 12-29-2024 08:18:22

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5389-02	1210-A	Water	pH	Cool 4 deg C	PSEG03	N31	12/27/2024	9040C
P5389-03	1210-B	Water	pH	Cool 4 deg C	PSEG03	N31	12/27/2024	9040C
P5389-05	1216-A	Water	pH	Cool 4 deg C	PSEG03	N31	12/27/2024	9040C
P5389-06	1216-B	Water	pH	Cool 4 deg C	PSEG03	N31	12/27/2024	9040C

Date/Time 12/29/24 08:25

Raw Sample Received by: JB WOC

Raw Sample Relinquished by: JB WOC

Date/Time 12/29/24 12:00

Raw Sample Received by: JB WOC

Raw Sample Relinquished by: JB WOC