

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB134119

Ambient Barometric Pressure (mmHg): 750.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3157

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	83.0	12/30/2024	09:05
2	P5386-05		1	13	100.00	>212.0	>212.0	12/30/2024	09:35
3	P5386-05DUP		1	13	100.00	>212.0	>212.0	12/30/2024	10:05
4	P5389-01		1	11	100.00	>212.0	>212.0	12/30/2024	10:35
5	P5389-04		1	11	100.00	>212.0	>212.0	12/30/2024	11:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66134119

WorkList Name : fp-12-27

WorkList ID : 186653

Department : Wet-Chemistry

Date : 12-27-2024 08:31:51

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5386-05 L	MOO-24-00397	Water	Flash Point	Cool 4 deg C	PSEG03	N31	12/26/2024	1010B
P5389-01 B	12-10-COMP	Water	Flash Point	Cool 4 deg C	PSEG03	N31	12/27/2024	1010B
P5389-04 B	1216-COMP	Water	Flash Point	Cool 4 deg C	PSEG03	N31	12/27/2024	1010B

Date/Time 12/30/2024 08:55
Raw Sample Received by: RM [Signature]
Raw Sample Relinquished by: CP (Jm)

Date/Time 12/30/2024 11:45
Raw Sample Received by: CP (Jm)
Raw Sample Relinquished by: RM [Signature]