


LB134120

7471B

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000
 B= 1.9055e-004
 C= -1.5708e-002
 Rho= 0.9999456
 Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/00
0.0	0.000	0.065	0.065	423	0.000	423					
0.2	0.200	0.188	-0.012	1070	0.0 %	1070					-6
2.5	2.500	2.446	-0.054	12919	0.0 %	12919					-2
5.0	5.000	4.967	-0.033	26152	0.0 %	26152					-1
7.5	7.500	7.517	0.017	39534	0.0 %	39534					0
10.0	10.000	10.016	0.016	52647	0.0 %	52647					0

LB134120 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 30 Dec 2024 11:27:00

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	423	-	0.0000	7471B	PPB	30 Dec 2024 11:36:36	S	Std
0.2 - 1	50.2	1070	-	0.2000	7471B	PPB	30 Dec 2024 11:38:53	S	Std
2.5 - 1	52.5	12919	-	2.5000	7471B	PPB	30 Dec 2024 11:41:09	S	Std
5.0 - 1	57	26152	-	5.0000	7471B	PPB	30 Dec 2024 11:43:26	S	Std
7.5 - 1	57.5	39534	-	7.5000	7471B	PPB	30 Dec 2024 11:45:45	S	Std
10.0 - 1	58	52647	-	10.0000	7471B	PPB	30 Dec 2024 11:50:47	S	Std
ICV37 - 1	ICV37	20153	3.8244	-	7471B	PPB	30 Dec 2024 11:53:31	U	SMPL
ICB37 - 1	ICB37	261	0.0340	-	7471B	PPB	30 Dec 2024 11:55:46	U	SMPL
CCV85 - 1	CCV85	26267	4.9894	-	7471B	PPB	30 Dec 2024 11:58:05	U	SMPL
CCB85 - 1	CCB85	180	0.0186	-	7471B	PPB	30 Dec 2024 12:00:20	U	SMPL
CRA - 1	CRA	1033	0.1811	-	7471B	PPB	30 Dec 2024 12:02:39	U	SMPL
HighStd - 1	HighStd	52942	10.0723	-	7471B	PPB	30 Dec 2024 12:04:54	U	SMPL
ChkStd - 1	ChkStd	36903	7.0161	-	7471B	PPB	30 Dec 2024 12:07:10	U	SMPL
PB165906BL - 1	PBS	-194	-0.0527	-	7471B	PPB	30 Dec 2024 12:09:32	U	SMPL
PB165906BS - 1	LCSS	21473	4.0759	-	7471B	PPB	30 Dec 2024 12:11:53	U	SMPL
P5386-01 - 1	MOO-24-00398	1020	0.1787	-	7471B	PPB	30 Dec 2024 12:14:10	U	SMPL
P5386-03 - 1	MOO-24-00395-96	602	0.0990	-	7471B	PPB	30 Dec 2024 12:16:29	U	SMPL
P5387-01 - 1	TR-05-122624	3369	0.6262	-	7471B	PPB	30 Dec 2024 12:18:47	U	SMPL
P5390-01 - 1	EO-01-12272024	3004	0.5567	-	7471B	PPB	30 Dec 2024 12:21:05	U	SMPL
CCV86 - 1	CCV86	26833	5.0973	-	7471B	PPB	30 Dec 2024 12:23:22	U	SMPL
CCB86 - 1	CCB86	296	0.0407	-	7471B	PPB	30 Dec 2024 12:25:38	U	SMPL
P5391-01 - 1	NB-07-12272024	807	0.1381	-	7471B	PPB	30 Dec 2024 12:27:57	U	SMPL
P5391-01DUP - 1	NB-07-12272024DUP	862	0.1485	-	7471B	PPB	30 Dec 2024 12:30:12	U	SMPL
P5391-01MS - 1	NB-07-12272024MS	21989	4.1742	-	7471B	PPB	30 Dec 2024 12:32:28	U	SMPL
P5391-01MSD - 1	NB-07-12272024MSD	20881	3.9631	-	7471B	PPB	30 Dec 2024 12:34:43	U	SMPL
P5391-01LX5 - 1		-66	-0.0283	-	7471B	PPB	30 Dec 2024 12:37:02	U	SMPL
P5391-01A - 1		21461	4.0736	-	7471B	PPB	30 Dec 2024 12:39:20	U	SMPL
CCV87 - 1	CCV87	26562	5.0456	-	7471B	PPB	30 Dec 2024 12:41:36	U	SMPL
CCB87 - 1	CCB87	313	0.0439	-	7471B	PPB	30 Dec 2024 12:43:56	U	SMPL