

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB134146

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3157

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	01/02/2025	14:45
2	P5402-01		1	10	100.00	>212.0	>212.0	01/02/2025	15:15
3	P5402-01DUP		1	10	100.00	>212.0	>212.0	01/02/2025	15:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB134446

WorkList Name : fp-1-02

WorkList ID : 186727

Department : Wet-Chemistry

Date : 01-02-2025 08:27:39

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5402-01	0827	Solid	Flash Point	Cool 4 deg C	PSEG03	L11	12/31/2024	1010B

Date/Time 01/02/2025 14:35
Raw Sample Received by: RM (wv)
Raw Sample Relinquished by: Mjwv

Date/Time 01/02/2025 16:25
Raw Sample Received by: Mjwv
Raw Sample Relinquished by: RMwv