

SFAM01.1


LB134147

INSTRUMENT ID: CV1

Linear

 μ Abs.:

50923



A= 0.0000e+000

B= 1.9840e-004 *slope*C= -2.4278e-002 *intercept*

Rho= 0.9998018

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.033	0.033	288	0.000	288	0				=
0.05	0.050					442	0				22
0.20	0.200	0.243	0.043	1349	0.0 %	1349					0
2.50	2.500	2.509	0.009	12769	0.0 %	12769					0
5.00	5.000	4.848	-0.152	24560	0.0 %	24560					3
7.50	7.500	7.488	-0.012	37864	0.0 %	37864					0
10.0	10.000	10.079	0.079	50923	0.0 %	50923					1

LB134147
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	288	-		0 SFAM01.1	PPB	1/2/2025 13:48	Std
	0.2 S01	1349	-		0.2 SFAM01.1	PPB	1/2/2025 13:50	Std
	2.5 S02	12769	-		2.5 SFAM01.1	PPB	1/2/2025 13:52	Std
	5 S03	24560	-		5 SFAM01.1	PPB	1/2/2025 13:55	Std
	7.5 S04	37864	-		7.5 SFAM01.1	PPB	1/2/2025 13:57	Std
	10 S05	50923	-		10 SFAM01.1	PPB	1/2/2025 14:06	Std
ICV040	ICV040	19947	3.9331 -		SFAM01.1	PPB	1/2/2025 14:12	SMPL
ICB040	ICB040	132	0.0019 -		SFAM01.1	PPB	1/2/2025 14:14	SMPL
CCV094	CCV094	24381	4.8128 -		SFAM01.1	PPB	1/2/2025 14:16	SMPL
CCB094	CCB094	50	-0.0144 -		SFAM01.1	PPB	1/2/2025 14:19	SMPL
PB165943BL	PBW943	59	-0.0126 -		SFAM01.1	PPB	1/2/2025 14:21	SMPL
P5347-01	MJNBZ9	233	0.0219 -		SFAM01.1	PPB	1/2/2025 14:23	SMPL
P5347-04	MJNC00	132	0.0019 -		SFAM01.1	PPB	1/2/2025 14:25	SMPL
P5347-08	MJNC00D	53	-0.0138 -		SFAM01.1	PPB	1/2/2025 14:28	SMPL
P5347-09	MJNC00S	4948	0.9574 -		SFAM01.1	PPB	1/2/2025 14:30	SMPL
P5354-01	ME2927	25	-0.0193 -		SFAM01.1	PPB	1/2/2025 14:32	SMPL
P5354-02	ME2928	46	-0.0152 -		SFAM01.1	PPB	1/2/2025 14:34	SMPL
P5354-03	ME2928D	42	-0.0159 -		SFAM01.1	PPB	1/2/2025 14:37	SMPL
P5354-04	ME2928S	4702	0.9086 -		SFAM01.1	PPB	1/2/2025 14:39	SMPL
CCV095	CCV095	23695	4.6767 -		SFAM01.1	PPB	1/2/2025 14:41	SMPL
CCB095	CCB095	98	-0.0048 -		SFAM01.1	PPB	1/2/2025 14:44	SMPL