

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB134153

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/03/2025	08:00
2	CAL2	1	Water	NA	NA	20.2	7.00	01/03/2025	08:01
3	CAL3	1	Water	NA	NA	20.2	10.02	01/03/2025	08:05
4	ICV	1	Water	NA	NA	20.1	7.00	01/03/2025	08:10
5	CCV1	1	Water	NA	NA	20.1	2.02	01/03/2025	08:30
6	Q1002-01	1	Solid	20.02	20	23.4	5.16	01/03/2025	08:37
7	Q1002-01DUP	1	Solid	20.03	20	23.5	5.17	01/03/2025	08:38
8	CCV2	1	Water	NA	NA	20.1	12.02	01/03/2025	08:40

WORKLIST(Hardcopy Internal Chain)

134153

WorkList Name : ph s q1002

WorkList ID : 186739

Department : Wet-Chemistry

Date : 01-03-2025 07:44:48

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1002-01	60051	Solid	pH	Cool 4 deg C	PSEG03	M11	01/02/2025	9045D

Date/Time 01-03-25 08:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 01-03-25 10:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]