

LB134155


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

51297



A= 0.0000e+000

B= 1.8851e-004

C= 5.4448e-003

Rho= 0.9980953

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9/01
0.0	0.000	0.011	0.011	31	0.000	31					10
0.2	0.200	0.220	0.020	1137	0.0 %	1137					-9
2.5	2.500	2.277	-0.223	12048	0.0 %	12048					3
5.0	5.000	5.155	0.155	27317	0.0 %	27317					5
7.5	7.500	7.862	0.362	41678	0.0 %	41678					-3
10.0	10.000	9.675	-0.325	51297	0.0 %	51297					

LB134155 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 03 Jan 2025 13:09:42

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	31	-	0.0000	7471B	PPB	03 Jan 2025 13:39:59	S	Std
0.2 - 1	50.2	1137	-	0.2000	7471B	PPB	03 Jan 2025 13:42:16	S	Std
2.5 - 1	52.5	12048	-	2.5000	7471B	PPB	03 Jan 2025 13:44:32	S	Std
5.0 - 1	55	27317	-	5.0000	7471B	PPB	03 Jan 2025 13:46:49	S	Std
7.5 - 1	57.5	41678	-	7.5000	7471B	PPB	03 Jan 2025 13:49:39	S	Std
10.0 - 1	510	51297	-	10.0000	7471B	PPB	03 Jan 2025 14:02:44	S	Std
ICV42 - 1	ICV42	20587	3.8863	-	7471B	PPB	03 Jan 2025 14:05:41	U	SMPL
ICB42 - 1	ICB42	-149	-0.0226	-	7471B	PPB	03 Jan 2025 14:07:56	U	SMPL
CCV01 - 1	CCV01	26777	5.0531	-	7471B	PPB	03 Jan 2025 14:12:59	U	SMPL
CCB01 - 1	CCB01	-215	-0.0351	-	7471B	PPB	03 Jan 2025 14:15:15	U	SMPL
CRA - 1	CRA	1122	0.2170	-	7471B	PPB	03 Jan 2025 14:17:32	U	SMPL
HighStd - 1	HighStd	53934	10.1724	-	7471B	PPB	03 Jan 2025 14:19:47	U	SMPL
ChkStd - 1	ChkStd	38481	7.2594	-	7471B	PPB	03 Jan 2025 14:22:02	U	SMPL
PB165952BL - 1	PBS	-334	-0.0575	-	7471B	PPB	03 Jan 2025 14:24:25	U	SMPL
PB165952BS - 1	LCSS	18029	3.4041	-	7471B	PPB	03 Jan 2025 14:26:46	U	SMPL
P5398-01 - 1	341-349	4984	0.9450	-	7471B	PPB	03 Jan 2025 14:34:01	U	SMPL
P5398-01DUP - 1	341-349DUP	4747	0.9003	-	7471B	PPB	03 Jan 2025 14:36:17	U	SMPL
P5398-01MS - 1	341-349MS	21865	4.1272	-	7471B	PPB	03 Jan 2025 14:38:35	U	SMPL
P5398-01MSD - 1	341-349MSD	21481	4.0548	-	7471B	PPB	03 Jan 2025 14:40:54	U	SMPL
CCV02 - 1	CCV02	26175	4.9396	-	7471B	PPB	03 Jan 2025 14:43:12	U	SMPL
CCB02 - 1	CCB02	-131	-0.0192	-	7471B	PPB	03 Jan 2025 14:45:30	U	SMPL
P5398-01LX5 - 1		1271	0.2450	-	7471B	PPB	03 Jan 2025 14:47:48	U	SMPL
P5398-01A - 1		21607	4.0785	-	7471B	PPB	03 Jan 2025 14:50:04	U	SMPL
CCV03 - 1	CCV03	26209	4.9461	-	7471B	PPB	03 Jan 2025 14:52:19	U	SMPL
CCB03 - 1	CCB03	-132	-0.0194	-	7471B	PPB	03 Jan 2025 14:54:37	U	SMPL