

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB134156

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/03/2025	14:10
2	CAL2	1	Water	NA	NA	20.3	7.00	01/03/2025	14:11
3	CAL3	1	Water	NA	NA	20.3	10.02	01/03/2025	14:15
4	ICV	1	Water	NA	NA	20.3	7.01	01/03/2025	14:20
5	CCV1	1	Water	NA	NA	20.2	2.01	01/03/2025	14:22
6	Q1008-02	1	Water	NA	NA	20.6	7.25	01/03/2025	14:30
7	Q1008-02DUP	1	Water	NA	NA	20.7	7.26	01/03/2025	14:31
8	Q1008-03	1	Water	NA	NA	20.1	5.64	01/03/2025	14:35
9	Q1008-04	1	Water	NA	NA	20.4	7.22	01/03/2025	14:40
10	Q1008-05	1	Water	NA	NA	20.3	6.09	01/03/2025	14:44
11	Q1008-07	1	Water	NA	NA	20.3	6.26	01/03/2025	14:47
12	Q1008-08	1	Water	NA	NA	20.7	7.19	01/03/2025	14:50
13	Q1008-09	1	Water	NA	NA	20.3	6.05	01/03/2025	15:00
14	Q1008-10	1	Water	NA	NA	20.1	7.15	01/03/2025	15:05
15	CCV2	1	Water	NA	NA	20.3	12.02	01/03/2025	15:10