

L8134165


7470A

INSTRUMENT ID: CV2

Linear

μ Abs.:

47528



A= 0.0000e+000

B= 2.1337e-004

C= 6.9369e-002

Rho= 0.9991618

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%
0.0	0.000	0.091	0.091	103	0.000	103					
0.2	0.200	0.201	0.001	615	0.0 %	615					
2.5	2.500	2.502	0.002	11399	0.0 %	11399					
5.0	5.000	4.986	-0.014	23044	0.0 %	23044					
7.5	7.500	7.210	-0.290	33466	0.0 %	33466					
10.0	10.000	10.210	0.210	47528	0.0 %	47528					

7
1
0
0
-4
2

LB134165 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 Jan 2025 11:32:05

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	103	-	0.0000	7470A	PPB	06 Jan 2025 11:45:17	S	Std
0.2 - 1	50-2	615	-	0.2000	7470A	PPB	06 Jan 2025 11:47:37	S	Std
2.5 - 1	12.5	11399	-	2.5000	7470A	PPB	06 Jan 2025 11:49:55	S	Std
5.0 - 1	55	23044	-	5.0000	7470A	PPB	06 Jan 2025 11:52:12	S	Std
7.5 - 1	57.5	33466	-	7.5000	7470A	PPB	06 Jan 2025 11:57:06	S	Std
10.0 - 1	510	47528	-	10.0000	7470A	PPB	06 Jan 2025 12:09:41	S	Std
ICV43 - 1	ICV43	18430	4.0017	-	7470A	PPB	06 Jan 2025 12:17:56	U	SMPL
ICB43 - 1	ICB43	-245	0.0171	-	7470A	PPB	06 Jan 2025 12:20:11	U	SMPL
CCV03 - 1	CCV03	25371	5.4827	-	7470A	PPB	06 Jan 2025 12:30:30	U	SMPL
CCB03 - 1	CCB03	-292	0.0071	-	7470A	PPB	06 Jan 2025 12:32:45	U	SMPL
CRA - 1	CRA	665	0.2113	-	7470A	PPB	06 Jan 2025 12:37:40	U	SMPL
HighStd - 1	HighStd	48630	10.4455	-	7470A	PPB	06 Jan 2025 12:39:55	U	SMPL
ChkStd - 1	ChkStd	30115	6.4950	-	7470A	PPB	06 Jan 2025 12:42:10	U	SMPL
PB165950BL - 1	PBW	-228	0.0207	-	7470A	PPB	06 Jan 2025 12:44:28	U	SMPL
PB165950BS - 1	LCSW	17466	3.7961	-	7470A	PPB	06 Jan 2025 12:46:46	U	SMPL
P5398-02 - 1	341-349	-45	0.0598	-	7470A	PPB	06 Jan 2025 12:49:02	U	SMPL
P5398-02DUP - 1	341-349DUP	138	0.0988	-	7470A	PPB	06 Jan 2025 12:53:59	U	SMPL
P5398-02MS - 1	341-349MS	18326	3.9796	-	7470A	PPB	06 Jan 2025 12:56:15	U	SMPL
P5398-02MSD - 1	341-349MSD	18446	4.0052	-	7470A	PPB	06 Jan 2025 12:58:33	U	SMPL
PB165951BL - 1	PBW	-204	0.0258	-	7470A	PPB	06 Jan 2025 13:00:49	U	SMPL
CCV04 - 1	CCV04	25365	5.4815	-	7470A	PPB	06 Jan 2025 13:03:07	U	SMPL
CCB04 - 1	CCB04	-243	0.0175	-	7470A	PPB	06 Jan 2025 13:05:22	U	SMPL
PB165951BS - 1	LCSW	18819	4.0847	-	7470A	PPB	06 Jan 2025 13:07:40	U	SMPL
P5386-05 - 1	MOO-24-00397	506	0.1773	-	7470A	PPB	06 Jan 2025 13:09:55	U	SMPL
P5386-05DUP - 1	MOO-24-00397DUP	701	0.2189	-	7470A	PPB	06 Jan 2025 13:12:12	U	SMPL
P5386-05MS - 1	MOO-24-00397MS	17318	3.7645	-	7470A	PPB	06 Jan 2025 13:14:29	U	SMPL
P5386-05MSD - 1	MOO-24-00397MSD	18870	4.0956	-	7470A	PPB	06 Jan 2025 13:16:47	U	SMPL
P5411-02 - 1	COMP	393	0.1532	-	7470A	PPB	06 Jan 2025 13:19:05	U	SMPL
P5398-02LX5 - 1		-20	0.0651	-	7470A	PPB	06 Jan 2025 13:27:02	U	SMPL
P5398-02A - 1		17297	3.7600	-	7470A	PPB	06 Jan 2025 13:29:18	U	SMPL
P5386-05LX5 - 1		-431	-0.0226	-	7470A	PPB	06 Jan 2025 13:31:35	U	SMPL
P5386-05A - 1		17084	3.7146	-	7470A	PPB	06 Jan 2025 13:33:52	U	SMPL
CCV05 - 1	CCV05	21168	4.5859	-	7470A	PPB	06 Jan 2025 13:36:08	U	SMPL
CCB05 - 1	CCB05	-168	0.0357	-	7470A	PPB	06 Jan 2025 13:38:26	U	SMPL
PB165934TB - 1	PB165934TB	82	0.0869	-	7470A	PPB	06 Jan 2025 13:40:43	U	SMPL
CCV06 - 1	CCV06	21180	4.5885	-	7470A	PPB	06 Jan 2025 13:42:58	U	SMPL
CCB06 - 1	CCB06	-154	0.0365	-	7470A	PPB	06 Jan 2025 13:45:14	U	SMPL