

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB134171

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3157

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	01/06/2025	14:00
2	Q1008-01		1	11	100.00	>212.0	>212.0	01/06/2025	14:30
3	Q1008-01DUP		1	11	100.00	>212.0	>212.0	01/06/2025	15:00
4	Q1008-06		1	14	100.00	>212.0	>212.0	01/06/2025	15:30
5	Q1008-10		1	12	100.00	>212.0	>212.0	01/06/2025	16:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6134171

WorkList Name : fp-1-6 WorkList ID : 186759 Department : Wet-Chemistry Date : 01-06-2025 08:38:42

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1008-01 A	COMP-1	Water	Flash Point	Cool 4 deg C	PSEG03	M11	01/03/2025	1010B
Q1008-06 A	COMP-2	Water	Flash Point	Cool 4 deg C	PSEG03	M11	01/03/2025	1010B
Q1008-10 C	DRUM-60414	Water	Flash Point	Cool 4 deg C	PSEG03	M11	01/03/2025	1010B

Date/Time 01/06/2025 13:50
 Raw Sample Received by: RM CWJ
 Raw Sample Relinquished by: RM CWJ

Date/Time 01/06/2025 16:40
 Raw Sample Received by: RM CWJ
 Raw Sample Relinquished by: RM CWJ