

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB134175

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3157

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	01/07/2025	09:40
2	Q1016-02		1	13	100.00	>212.0	>212.0	01/07/2025	10:10
3	Q1016-02DUP		1	13	100.00	>212.0	>212.0	01/07/2025	10:40
4	Q1018-05		1	11	100.00	>212.0	>212.0	01/07/2025	11:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB134175

WorkList Name : fp-1-7

WorkList ID : 186787

Department : Wet-Chemistry

Date : 01-07-2025 08:21:55

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1016-02	275	Water	Flash Point	Cool 4 deg C	PSEG03	N41	01/06/2025	1010B
Q1018-05	3730	Water	Flash Point	Cool 4 deg C	PSEG03	N41	01/06/2025	1010B

Date/Time 01/07/2025 09:30
Raw Sample Received by: R17 (wlc)
Raw Sample Relinquished by: JH (wlc)

Date/Time 01/07/2025 11:50
Raw Sample Received by: JH (wlc)
Raw Sample Relinquished by: R17 (wlc)