

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB134179

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/07/2025	08:30
2	CAL2	1	Water	NA	NA	20.1	7.00	01/07/2025	08:31
3	CAL3	1	Water	NA	NA	20.1	10.03	01/07/2025	08:33
4	ICV	1	Water	NA	NA	20.3	6.99	01/07/2025	08:35
5	CCV1	1	Water	NA	NA	20.2	2.01	01/07/2025	08:36
6	Q1016-02	1	Water	NA	NA	20.6	7.40	01/07/2025	08:44
7	Q1016-02DUP	1	Water	NA	NA	20.7	7.41	01/07/2025	08:45
8	Q1018-05	1	Water	NA	NA	20.6	7.70	01/07/2025	08:50
9	CCV2	1	Water	NA	NA	20.2	12.02	01/07/2025	08:55

WORKLIST(Hardcopy Internal Chain)

VB134149

WorkList Name : ph q1018 water

WorkList ID : 186783

Department : Wet-Chemistry

Date : 01-07-2025 08:14:27

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1016-02	275	Water	pH	Cool 4 deg C	PSEG03	N41	01/06/2025	9040C
Q1018-05	3730	Water	pH	Cool 4 deg C	PSEG03	N41	01/06/2025	9040C

Date/Time 01/07/25 08:25

Raw Sample Received by: JB we

Raw Sample Relinquished by: JB we

Date/Time 01/07/25

Raw Sample Received by: JB we

Raw Sample Relinquished by: JB we

11:40

JB we

JB we