

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB134180

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/07/2025	08:30
2	CAL2	1	Water	NA	NA	20.1	7.00	01/07/2025	08:31
3	CAL3	1	Water	NA	NA	20.1	10.03	01/07/2025	08:33
4	ICV	1	Water	NA	NA	20.3	6.99	01/07/2025	08:35
5	CCV1	1	Water	NA	NA	20.1	2.01	01/07/2025	09:10
6	Q1016-01	1	Solid	20.02	20	20.1	5.65	01/07/2025	09:17
7	Q1018-01	1	Solid	20.03	20	20.1	5.32	01/07/2025	09:20
8	Q1018-01DUP	1	Solid	20.04	20	20.2	5.33	01/07/2025	09:21
9	Q1018-02	1	Solid	20.01	20	20.6	5.07	01/07/2025	09:24
10	Q1018-03	1	Solid	20.03	20	20.2	6.40	01/07/2025	09:29
11	Q1018-04	1	Solid	20.02	20	20.2	6.26	01/07/2025	09:35
12	Q1021-01	1	Solid	20.02	20	20.2	6.26	01/07/2025	09:40
13	CCV2	1	Water	NA	NA	20.2	12.02	01/07/2025	09:44

WORKLIST(Hardcopy Internal Chain)

VB 134180

WorkList Name : corrosivity q1018 s WorkList ID : 186784 Department : Wet-Chemistry Date : 01-07-2025 08:15:07

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1016-01	276	Solid	Corrosivity	Cool 4 deg C	PSEG03	N41	01/06/2025	9045D
Q1018-01	3732	Solid	Corrosivity	Cool 4 deg C	PSEG03	N41	01/06/2025	9045D
Q1018-02	3735	Solid	Corrosivity	Cool 4 deg C	PSEG03	N41	01/06/2025	9045D
Q1018-03	3744	Solid	Corrosivity	Cool 4 deg C	PSEG03	N41	01/06/2025	9045D
Q1018-04	3745	Solid	Corrosivity	Cool 4 deg C	PSEG03	N41	01/06/2025	9045D
Q1021-01	424	Solid	Corrosivity	Cool 4 deg C	PSEG03	N41	01/06/2025	9045D

Date/Time 01/07/25 08:20

Raw Sample Received by: JP Wong

Raw Sample Relinquished by: JP Wong

Date/Time 01/07/25

Raw Sample Received by: JP Wong

Raw Sample Relinquished by: JP Wong

12:00

JP Wong

JP Wong