


LB134182

7471B

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.0175e-004

C= 3.3941e-002

Rho= 0.9984623

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0.0
0.0	0.000	0.047	0.047	63	0.000	63					
0.2	0.200	0.213	0.013	887	0.0 %	887					6
2.5	2.500	2.405	-0.095	11752	0.0 %	11752					-4
5.0	5.000	4.871	-0.129	23977	0.0 %	23977					-3
7.5	7.500	7.906	0.406	39018	0.0 %	39018					5
10.0	10.000	9.758	-0.242	48200	0.0 %	48200					-2

LB134182 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 07 Jan 2025 13:42:34

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	63	-	0.000	7471B	PPB	07 Jan 2025 13:45:00	S	Std
0.2 - 1	50.2	887	-	0.200	7471B	PPB	07 Jan 2025 13:47:16	S	Std
2.5 - 1	52.5	11752	-	2.500	7471B	PPB	07 Jan 2025 13:49:34	S	Std
5.0 - 1	55	23977	-	5.000	7471B	PPB	07 Jan 2025 13:51:51	S	Std
7.5 - 1	57.5	39018	-	7.500	7471B	PPB	07 Jan 2025 13:54:09	S	Std
10.0 - 1	59	48200	-	10.000	7471B	PPB	07 Jan 2025 14:04:39	S	Std
ICV47 - 1	ICV47	17766	3.6183	-	7471B	PPB	07 Jan 2025 14:07:30	U	SMPL
ICB47 - 1	ICB47	-238	-0.0141	-	7471B	PPB	07 Jan 2025 14:09:45	U	SMPL
CCV11 - 1	CCV11	25470	5.1725	-	7471B	PPB	07 Jan 2025 14:12:03	U	SMPL
CCB11 - 1	CCB11	-276	-0.0217	-	7471B	PPB	07 Jan 2025 14:14:18	U	SMPL
CRA - 1	CRA	874	0.2103	-	7471B	PPB	07 Jan 2025 14:16:37	U	SMPL
HighStd - 1	HighStd	48274	9.7733	-	7471B	PPB	07 Jan 2025 14:18:52	U	SMPL
ChkStd - 1	ChkStd	36101	7.3174	-	7471B	PPB	07 Jan 2025 14:21:07	U	SMPL
PB165972BL - 1	PBS	-367	-0.0401	-	7471B	PPB	07 Jan 2025 14:23:30	U	SMPL
PB165972BS - 1	LCSS	18079	3.6814	-	7471B	PPB	07 Jan 2025 14:25:50	U	SMPL
Q1007-01 - 1	WD-01	18063	3.6782	-	7471B	PPB	07 Jan 2025 14:28:07	U	SMPL
Q1009-01 - 1	SU-03-1-03-2025	1260	0.2881	-	7471B	PPB	07 Jan 2025 14:30:26	U	SMPL
Q1009-01MS - 1	SU-03-1-03-2025MS	24308	4.9381	-	7471B	PPB	07 Jan 2025 14:35:00	U	SMPL
Q1009-01MSD - 1	SU-03-1-03-2025MSD	21988	4.4700	-	7471B	PPB	07 Jan 2025 14:37:14	U	SMPL
CCV12 - 1	CCV12	25276	5.1334	-	7471B	PPB	07 Jan 2025 14:39:33	U	SMPL
CCB12 - 1	CCB12	-180	-0.0024	-	7471B	PPB	07 Jan 2025 14:41:51	U	SMPL
Q1010-01 - 1	TR-04-010325	1899	0.4171	-	7471B	PPB	07 Jan 2025 14:44:09	U	SMPL
Q1016-01 - 1	276	26	0.0392	-	7471B	PPB	07 Jan 2025 14:46:25	U	SMPL
Q1009-01A - 1		20808	4.2320	-	7471B	PPB	07 Jan 2025 14:50:56	U	SMPL
Q1009-01DUP - 1		1399	0.3162	-	7471B	PPB	07 Jan 2025 14:53:12	U	SMPL
Q1009-01LX5 - 1		167	0.0676	-	7471B	PPB	07 Jan 2025 15:00:54	U	SMPL
CCV13 - 1	CCV13	25773	5.2337	-	7471B	PPB	07 Jan 2025 15:03:28	U	SMPL
CCB13 - 1	CCB13	-176	-0.0016	-	7471B	PPB	07 Jan 2025 15:05:44	U	SMPL