

LB134193


7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

37263



A= 0.0000e+000

B= 2.6665e-004

C= -8.7389e-002

Rho= 0.9995433

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.069	-0.069	69	0.000	69	0			
0.2	0.200	0.161	-0.039	930	0.0 %	930				
2.5	2.500	2.498	-0.002	9694	0.0 %	9694				
5.0	5.000	5.206	0.206	19850	0.0 %	19850				
7.5	7.500	7.556	0.056	28665	0.0 %	28665				
10.0	10.000	9.849	-0.151	37263	0.0 %	37263				

✓
1
20
0
4
1
2

LB134193 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 08 Jan 2025 13:21:51

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	69	-	0.0000	7470A	PPB	08 Jan 2025 13:24:06	S	Std
0.2 - 1	50.2	930	-	0.2000	7470A	PPB	08 Jan 2025 13:26:24	S	Std
2.5 - 1	52.5	9694	-	2.5000	7470A	PPB	08 Jan 2025 13:28:41	S	Std
5.0 - 1	85	19850	-	5.0000	7470A	PPB	08 Jan 2025 13:30:57	S	Std
7.5 - 1	87.5	28665	-	7.5000	7470A	PPB	08 Jan 2025 13:33:14	S	Std
10.0 - 1	516	37263	-	10.0000	7470A	PPB	08 Jan 2025 13:43:49	S	Std
ICV49 - 1	ICV49	14947	3.8983	-	7470A	PPB	08 Jan 2025 13:46:41	U	SMPL
ICB49 - 1	ICB49	35	-0.0781	-	7470A	PPB	08 Jan 2025 13:51:36	U	SMPL
CCV17 - 1	CCV17	20252	5.3129	-	7470A	PPB	08 Jan 2025 13:53:51	U	SMPL
CCB17 - 1	CCB17	154	-0.0463	-	7470A	PPB	08 Jan 2025 13:58:47	U	SMPL
CRA - 1	CRA	1134	0.2150	-	7470A	PPB	08 Jan 2025 14:09:05	U	SMPL
HighStd - 1	HighStd	37400	9.8855	-	7470A	PPB	08 Jan 2025 14:11:20	U	SMPL
ChkStd - 1	ChkStd	27882	7.3474	-	7470A	PPB	08 Jan 2025 14:13:35	U	SMPL
PB165982BL - 1	PBW	129	-0.0530	-	7470A	PPB	08 Jan 2025 14:18:29	U	SMPL
PB165982BS - 1	LCSW	17677	4.6262	-	7470A	PPB	08 Jan 2025 14:20:46	U	SMPL
Q1018-01 - 1	3732	-77	-0.1079	-	7470A	PPB	08 Jan 2025 14:23:02	U	SMPL
Q1018-01DUP - 1	3732DUP	101	-0.0605	-	7470A	PPB	08 Jan 2025 14:25:21	U	SMPL
Q1018-01MS - 1	3732MS	15412	4.0223	-	7470A	PPB	08 Jan 2025 14:32:39	U	SMPL
Q1018-01MSD - 1	3732MSD	17147	4.4849	-	7470A	PPB	08 Jan 2025 14:34:56	U	SMPL
Q1018-02 - 1	3735	-209	-0.1431	-	7470A	PPB	08 Jan 2025 14:37:12	U	SMPL
CCV18 - 1	CCV18	20581	5.4006	-	7470A	PPB	08 Jan 2025 14:39:29	U	SMPL
CCB18 - 1	CCB18	13	-0.0839	-	7470A	PPB	08 Jan 2025 14:41:44	U	SMPL
Q1018-03 - 1	3744	46	-0.0751	-	7470A	PPB	08 Jan 2025 14:44:02	U	SMPL
Q1018-04 - 1	3745	75	-0.0674	-	7470A	PPB	08 Jan 2025 14:46:17	U	SMPL
Q1021-01 - 1	424	93	-0.0626	-	7470A	PPB	08 Jan 2025 14:48:32	U	SMPL
PB165973TB - 1	PB165973TB	107	-0.0589	-	7470A	PPB	08 Jan 2025 14:50:48	U	SMPL
Q1018-01LX5 - 1		23	-0.0813	-	7470A	PPB	08 Jan 2025 14:53:04	U	SMPL
Q1018-01A - 1		15300	3.9924	-	7470A	PPB	08 Jan 2025 14:55:20	U	SMPL
CCV19 - 1	CCV19	18854	4.9401	-	7470A	PPB	08 Jan 2025 15:00:32	U	SMPL
CCB19 - 1	CCB19	57	-0.0722	-	7470A	PPB	08 Jan 2025 15:02:49	U	SMPL