

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB134209

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3157

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	01/09/2025	09:50
2	Q1032-01		1	14	100.00	>212.0	>212.0	01/09/2025	10:20
3	Q1032-01DUP		1	14	100.00	>212.0	>212.0	01/09/2025	10:50
4	Q1032-04		1	12	100.00	>212.0	>212.0	01/09/2025	11:20
5	Q1032-07		1	11	100.00	>212.0	>212.0	01/09/2025	11:50
6	Q1033-01		1	14	100.00	>212.0	>212.0	01/09/2025	12:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6134209

WorkList Name : fp-1-8 WorkList ID : 186813 Department : Wet-Chemistry Date : 01-08-2025 13:27:08

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1032-01 F	50492	Water	Flash Point	Cool 4 deg C	PSEG03	N41	01/08/2025	1010B
Q1032-04	0103-COMP	Water	Flash Point	Cool 4 deg C	PSEG03	N41	01/08/2025	1010B
Q1032-07 C	0107-COMP	Water	Flash Point	Cool 4 deg C	PSEG03	N41	01/08/2025	1010B
Q1033-01 O	28612	Water	Flash Point	Cool 4 deg C	PSEG03	N41	01/08/2025	1010B

Date/Time 01/09/2025 0940
 Raw Sample Received by: RM (WC)
 Raw Sample Relinquished by: RM (WC)

Date/Time 01/09/2025 1300
 Raw Sample Received by: RM (WC)
 Raw Sample Relinquished by: RM (WC)