

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB134226

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3157

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	01/10/2025	11:10
2	Q1049-01		1	14	100.00	>212.0	>212.0	01/10/2025	11:40
3	Q1052-01		1	11	100.00	>212.0	>212.0	01/10/2025	12:25
4	Q1052-01DUP		1	11	100.00	>212.0	>212.0	01/10/2025	12:55

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LD134226

WorkList Name : FP-1-9

WorkList ID : 186848

Department : Wet-Chemistry

Date : 01-09-2025 16:31:53

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1049-01	FRAC-TANK-257952	Water	Flash Point	Cool 4 deg C	PSEG03	N41	01/09/2025	1010B
Q1052-01	112124	Water	Flash Point	Cool 4 deg C	PSEG03	N51	01/09/2025	1010B

Date/Time 01/10/2025 11.00
Raw Sample Received by: RM WWC
Raw Sample Relinquished by: RM WWC

Date/Time 01/10/2025 13.35
Raw Sample Received by: RM WWC
Raw Sample Relinquished by: RM WWC