

LB134231


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

39360



A= 0.0000e+000

B= 2.5339e-004

C= 6.6507e-002

Rho= 0.9997790

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/1)
0.0	0.000	0.005	0.005	-241	0.000	-241	0				-
0.2	0.200	0.257	0.057	750	0.0 %	750					29
2.5	2.500	2.386	-0.114	9153	0.0 %	9153					-5
5.0	5.000	5.101	0.101	19870	0.0 %	19870					2
7.5	7.500	7.411	-0.089	28984	0.0 %	28984					-1
10.0	10.000	10.040	0.040	39360	0.0 %	39360					0

LB134231 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 10 Jan 2025 14:35:12

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-241	-	0.0000	7470A	PPB	10 Jan 2025 15:14:35	S	Std
0.2 - 1	50.2	750	-	0.2000	7470A	PPB	10 Jan 2025 15:16:53	S	Std
2.5 - 1	52.5	9153	-	2.5000	7470A	PPB	10 Jan 2025 15:19:11	S	Std
5.0 - 1	55	19870	-	5.0000	7470A	PPB	10 Jan 2025 15:21:28	S	Std
7.5 - 1	57.5	28984	-	7.5000	7470A	PPB	10 Jan 2025 15:28:57	S	Std
10.0 - 1	59	39360	-	10.0000	7470A	PPB	10 Jan 2025 15:33:53	S	Std
ICV52 - 1	ICV52	15781	4.0653	-	7470A	PPB	10 Jan 2025 15:36:45	U	SMPL
ICB52 - 1	ICB52	-90	0.0437	-	7470A	PPB	10 Jan 2025 15:39:00	U	SMPL
CCV25 - 1	CCV25	20135	5.1685	-	7470A	PPB	10 Jan 2025 15:41:17	U	SMPL
CCB25 - 1	CCB25	-45	0.0551	-	7470A	PPB	10 Jan 2025 15:43:32	U	SMPL
CRA - 1	CRA	735	0.2527	-	7470A	PPB	10 Jan 2025 15:48:20	U	SMPL
HighStd - 1	HighStd	39557	10.0899	-	7470A	PPB	10 Jan 2025 15:50:35	U	SMPL
ChkStd - 1	ChkStd	27056	6.9223	-	7470A	PPB	10 Jan 2025 15:52:50	U	SMPL
PB166024BL - 1	PBW	-55	0.0526	-	7470A	PPB	10 Jan 2025 15:55:08	U	SMPL
PB166024BS - 1	LCSW	14838	3.8263	-	7470A	PPB	10 Jan 2025 15:57:26	U	SMPL
Q1049-01 - 1	FRAC-TANK-257952	2443	0.6855	-	7470A	PPB	10 Jan 2025 15:59:43	U	SMPL
Q1049-01DUP - 1	FRAC-TANK-257952DUP	2789	0.7732	-	7470A	PPB	10 Jan 2025 16:02:01	U	SMPL
Q1049-01MS - 1	FRAC-TANK-257952MS	14422	3.7209	-	7470A	PPB	10 Jan 2025 16:07:08	U	SMPL
Q1049-01MSD - 1	FRAC-TANK-257952MSD	14505	3.7420	-	7470A	PPB	10 Jan 2025 16:09:24	U	SMPL
CCV26 - 1	CCV26	19203	4.9324	-	7470A	PPB	10 Jan 2025 16:14:19	U	SMPL
CCB26 - 1	CCB26	-99	0.0414	-	7470A	PPB	10 Jan 2025 16:16:33	U	SMPL
Q1049-01LX5 - 1		455	0.1818	-	7470A	PPB	10 Jan 2025 16:18:50	U	SMPL
Q1049-01A - 1		15412	3.9718	-	7470A	PPB	10 Jan 2025 16:21:05	U	SMPL
CCV27 - 1	CCV27	20433	5.2441	-	7470A	PPB	10 Jan 2025 16:23:21	U	SMPL
CCB27 - 1	CCB27	-18	0.0619	-	7470A	PPB	10 Jan 2025 16:25:38	U	SMPL