

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB134234

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/10/2025	16:10
2	CAL2	1	Water	NA	NA	20.1	7.00	01/10/2025	16:11
3	CAL3	1	Water	NA	NA	20.2	10.02	01/10/2025	16:15
4	ICV	1	Water	NA	NA	20.1	6.99	01/10/2025	16:20
5	CCV1	1	Water	NA	NA	20.2	2.01	01/10/2025	16:22
6	Q1049-01	1	Water	NA	NA	20.7	8.50	01/10/2025	16:30
7	Q1049-01DUP	1	Water	NA	NA	20.8	8.51	01/10/2025	16:31
8	Q1050-01	1	Water	NA	NA	20.1	3.96	01/10/2025	16:37
9	Q1052-01	1	Water	NA	NA	20.6	8.45	01/10/2025	16:44
10	CCV2	1	Water	NA	NA	20.3	12.02	01/10/2025	16:45

WORKLIST(Hardcopy Internal Chain)

134234

WorkList Name : ph q1049

WorkList ID : 186874

Department : Wet-Chemistry

Date : 01-10-2025 15:47:16

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1049-01	FRAC-TANK-257952	Water	pH	Cool 4 deg C	PSEG03	N41	01/09/2025	9040C
Q1050-01	COND-TOTE	Water	pH	Cool 4 deg C	PSEG03	N51	01/09/2025	9040C
Q1052-01	112124	Water	pH	Cool 4 deg C	PSEG03	N51	01/09/2025	9040C

Date/Time 01-10-25 15:50

Raw Sample Received by: 28 WOC

Raw Sample Relinquished by: CP SM

Date/Time 01-10-25

Raw Sample Received by: CP SM

Raw Sample Relinquished by: 28 WOC