

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB134243

Slope : 98.4

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/13/2025	09:00
2	CAL2	1	Water	NA	NA	20.2	7.00	01/13/2025	09:01
3	CAL3	1	Water	NA	NA	20.2	10.02	01/13/2025	09:05
4	ICV	1	Water	NA	NA	20.2	7.02	01/13/2025	09:10
5	CCV1	1	Water	NA	NA	20.1	2.01	01/13/2025	09:11
6	Q1069-01	1	Solid	20.02	20	20.7	5.57	01/13/2025	09:20
7	Q1069-01DUP	1	Solid	20.03	20	20.2	5.58	01/13/2025	09:21
8	CCV2	1	Water	NA	NA	20.2	12.02	01/13/2025	09:25

WORKLIST(Hardcopy Internal Chain)

VB 134243

WorkList Name : ph q1069 WorkList ID : 186886 Department : Wet-Chemistry Date : 01-13-2025 08:46:32

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1069-01	RW7B-CARBON-20250109	Solid	pH	Cool 4 deg C	TETRO6	M11	01/09/2025	9045D

Date/Time 01-13-25 08:50
Raw Sample Received by: JB WLC
Raw Sample Relinquished by: CCSR

Date/Time 01-13-25 11:30
Raw Sample Received by: CCSR
Raw Sample Relinquished by: JB CCR