

LB134280

7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

44602

A= 0.0000e+000

B= 2.1798e-004

C= -4.3511e-002

Rho= 0.9982029

Accept=Accepted

Max. Conc: 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/1
0.0	0.000	-0.093	-0.093	-229	0.000	-229					
0.2	0.200	0.247	0.047	1333	0.0 %	1333					24
2.5	2.500	2.493	-0.007	11635	0.0 %	11635					0
5.0	5.000	4.954	-0.046	22925	0.0 %	22925					-1
7.5	7.500	7.921	0.421	36537	0.0 %	36537					6
10.0	10.000	9.679	-0.321	44602	0.0 %	44602					-3

LB134280 INSTRUMENT ID : C1

Method: 7471B

Operator: Admin

Date of Analysis: 14 Jan 2025 10:53:06

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-229	-	0.0000	7471B	PPB	14 Jan 2025 11:08:00	S	Std
0.2 - 1	50.2	1333	-	0.2000	7471B	PPB	14 Jan 2025 11:10:16	S	Std
2.5 - 1	52.5	11635	-	2.5000	7471B	PPB	14 Jan 2025 11:12:33	S	Std
5.0 - 1	55	22925	-	5.0000	7471B	PPB	14 Jan 2025 11:14:50	S	Std
7.5 - 1	57.5	36537	-	7.5000	7471B	PPB	14 Jan 2025 11:17:08	S	Std
10.0 - 1	59	44602	-	10.0000	7471B	PPB	14 Jan 2025 11:22:16	S	Std
ICV54 - 1	ICV54	19159	4.1328	-	7471B	PPB	14 Jan 2025 11:25:10	U	SMPL
ICB54 - 1	ICB54	-53	-0.0551	-	7471B	PPB	14 Jan 2025 11:27:27	U	SMPL
CCV29 - 1	CCV29	23002	4.9705	-	7471B	PPB	14 Jan 2025 11:29:45	U	SMPL
CCB29 - 1	CCB29	-76	-0.0601	-	7471B	PPB	14 Jan 2025 11:32:00	U	SMPL
CRA - 1	CRA	1187	0.2152	-	7471B	PPB	14 Jan 2025 11:36:51	U	SMPL
HighStd - 1	HighStd	44485	9.6535	-	7471B	PPB	14 Jan 2025 11:39:06	U	SMPL
ChkStd - 1	ChkStd	34111	7.3921	-	7471B	PPB	14 Jan 2025 11:41:21	U	SMPL
PB166048BL - 1	PBS	-235	-0.0947	-	7471B	PPB	14 Jan 2025 11:43:46	U	SMPL
PB166048BS - 1	LCSS	17833	3.8438	-	7471B	PPB	14 Jan 2025 11:46:07	U	SMPL
Q1068-01 - 1	TR-06-1-10-2025	1653	0.3168	-	7471B	PPB	14 Jan 2025 11:53:23	U	SMPL
Q1068-01DUP - 1	TR-06-1-10-2025DUP	1546	0.2935	-	7471B	PPB	14 Jan 2025 11:55:39	U	SMPL
Q1068-01MS - 1	TR-06-1-10-2025MS	22205	4.7968	-	7471B	PPB	14 Jan 2025 11:57:56	U	SMPL
Q1068-01MSD - 1	TR-06-1-10-2025MSD	22483	4.8574	-	7471B	PPB	14 Jan 2025 12:00:13	U	SMPL
Q1074-01 - 1	HACKENSACK-DGA	823	0.1359	-	7471B	PPB	14 Jan 2025 12:02:31	U	SMPL
CCV30 - 1	CCV30	22781	4.9224	-	7471B	PPB	14 Jan 2025 12:04:50	U	SMPL
CCB30 - 1	CCB30	-113	-0.0681	-	7471B	PPB	14 Jan 2025 12:07:05	U	SMPL
Q1076-01 - 1	ARS20-0006	1865	0.3630	-	7471B	PPB	14 Jan 2025 12:09:23	U	SMPL
Q1077-01 - 1	TR-05-01132025	1786	0.3458	-	7471B	PPB	14 Jan 2025 12:11:39	U	SMPL
Q1068-01LX5 - 1		201	0.0003	-	7471B	PPB	14 Jan 2025 12:13:56	U	SMPL
Q1068-01A - 1		21211	4.5801	-	7471B	PPB	14 Jan 2025 12:16:12	U	SMPL
CCV31 - 1	CCV31	22615	4.8862	-	7471B	PPB	14 Jan 2025 12:18:28	U	SMPL
CCB31 - 1	CCB31	9	-0.0415	-	7471B	PPB	14 Jan 2025 12:20:47	U	SMPL