

LB134281


7470A

INSTRUMENT ID: CY1

Linear

 μ Abs.:

41197



A= 0.0000e+000

B= 2.4141e-004

C= -1.3931e-001

Rho= 0.9990659

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.014	0.014	634	0.000	634				
0.2	0.200	0.240	0.040	1571	0.0 %	1571	0			
2.5	2.500	2.379	-0.121	10432	0.0 %	10432				
5.0	5.000	4.951	-0.049	21086	0.0 %	21086				
7.5	7.500	7.810	0.310	32929	0.0 %	32929				
10.0	10.000	9.806	-0.194	41197	0.0 %	41197				

0/017
 1
 20
 -5
 1
 4
 2

LB134281 INSTRUMENT ID : CV1

Method:7470A Operator : Mohan

Date Of Analysis : 14 january 2025 13:31:00

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0	634	-		0 7470A	PPB	1/14/2025 13:38	Std
	0.2	1571	-		0.2 7470A	PPB	1/14/2025 13:40	Std
	2.5	10432	-		2.5 7470A	PPB	1/14/2025 13:42	Std
	5	21086	-		5 7470A	PPB	1/14/2025 13:45	Std
	7.5	32929	-		7.5 7470A	PPB	1/14/2025 13:52	Std
	10	41197	-		10 7470A	PPB	1/14/2025 13:57	Std
ICV55	ICV55	15842	3.6851 -		7470A	PPB	1/14/2025 14:00	SMPL
ICB55	ICB55	412	-0.0399 -		7470A	PPB	1/14/2025 14:02	SMPL
CCV32	CCV32	21245	4.9895 -		7470A	PPB	1/14/2025 14:07	SMPL
CCB32	CCB32	494	-0.0201 -		7470A	PPB	1/14/2025 14:09	SMPL
CRA	CRA	1448	0.2103 -		7470A	PPB	1/14/2025 14:14	SMPL
HighStd	HighStd	40086	9.5379 -		7470A	PPB	1/14/2025 14:16	SMPL
ChkStd	ChkStd	29320	6.9388 -		7470A	PPB	1/14/2025 14:19	SMPL
PB166050BL	PBW	658	0.0195 -		7470A	PPB	1/14/2025 14:21	SMPL
PB166050BS	LCSW	16009	3.7254 -		7470A	PPB	1/14/2025 14:23	SMPL
Q1036-02	NP-WS-002	912	0.0809 -		7470A	PPB	1/14/2025 14:26	SMPL
Q1036-02DUP	NP-WS-002DUP	848	0.0654 -		7470A	PPB	1/14/2025 14:28	SMPL
Q1036-02MS	NP-WS-002MS	16513	3.8471 -		7470A	PPB	1/14/2025 14:30	SMPL
Q1036-02MSD	NP-WS-002MSD	15749	3.6627 -		7470A	PPB	1/14/2025 14:32	SMPL
PB166051BL	PBW	857	0.0676 -		7470A	PPB	1/14/2025 14:35	SMPL
CCV33	CCV33	20915	4.9098 -		7470A	PPB	1/14/2025 14:37	SMPL
CCB33	CCB33	371	-0.0497 -		7470A	PPB	1/14/2025 14:39	SMPL
PB166051BS	LCSW	15676	3.645 -		7470A	PPB	1/14/2025 14:42	SMPL
Q1065-01	MOO-25-0002	-234	-0.1958 -		7470A	PPB	1/14/2025 14:44	SMPL
Q1069-02	RW7B-CARBON-20250109	965	0.0936 -		7470A	PPB	1/14/2025 14:46	SMPL
Q1074-02	HACKENSACK-DGA	523	-0.0131 -		7470A	PPB	1/14/2025 14:48	SMPL
Q1076-02	ARS20-0006	745	0.0405 -		7470A	PPB	1/14/2025 14:58	SMPL
Q1076-02DUP	ARS20-0006DUP	778	0.0485 -		7470A	PPB	1/14/2025 15:00	SMPL
Q1076-02MS	ARS20-0006MS	17256	4.0265 -		7470A	PPB	1/14/2025 15:02	SMPL
Q1076-02MSD	ARS20-0006MSD	17222	4.0183 -		7470A	PPB	1/14/2025 15:05	SMPL
PB166025TB	PB166025TB	737	0.0386 -		7470A	PPB	1/14/2025 15:07	SMPL
PB166026TB	PB166026TB	931	0.0854 -		7470A	PPB	1/14/2025 15:09	SMPL
CCV34	CCV34	20017	4.693 -		7470A	PPB	1/14/2025 15:12	SMPL
CCB34	CCB34	441	-0.0328 -		7470A	PPB	1/14/2025 15:14	SMPL
Q1036-02LX5		322	-0.0616 -		7470A	PPB	1/14/2025 15:16	SMPL
Q1036-02A		14699	3.4092 -		7470A	PPB	1/14/2025 15:18	SMPL
Q1076-02LX5		69	-0.1227 -		7470A	PPB	1/14/2025 15:21	SMPL
Q1076-02A		16853	3.9292 -		7470A	PPB	1/14/2025 15:23	SMPL
CCV35	CCV35	19447	4.5554 -		7470A	PPB	1/14/2025 15:25	SMPL
CCB35	CCB35	340	-0.0572 -		7470A	PPB	1/14/2025 15:28	SMPL