

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB134298

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3157

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	01/16/2025	09:10
2	Q1092-01		1	11	100.00	>212.0	>212.0	01/16/2025	09:40
3	Q1092-01DUP		1	11	100.00	>212.0	>212.0	01/16/2025	10:10
4	Q1094-03		1	14	100.00	>212.0	>212.0	01/16/2025	10:40
5	Q1096-01		1	14	100.00	>212.0	>212.0	01/16/2025	11:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6134298

WorkList Name : FP-1-15 WorkList ID : 186920 Department : Wet-Chemistry Date : 01-15-2025 16:20:44

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1092-01 D	427	Water	Flash Point	Cool 4 deg C	PSEG03	N41	01/15/2025	1010B
Q1094-03 D	3181	Water	Flash Point	Cool 4 deg C	PSEG03	N41	01/15/2025	1010B
Q1096-01 D	D3746	Water	Flash Point	Cool 4 deg C	PSEG03	N41	01/15/2025	1010B

Date/Time 01/16/2025 08:55
 Raw Sample Received by: RM WLC
 Raw Sample Relinquished by: WLC

Date/Time 01/16/2025 11:45
 Raw Sample Received by: WLC
 Raw Sample Relinquished by: RM WLC