

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB134300

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/16/2025	08:15
2	CAL2	1	Water	NA	NA	20.2	7.01	01/16/2025	08:16
3	CAL3	1	Water	NA	NA	20.2	10.02	01/16/2025	08:20
4	ICV	1	Water	NA	NA	20.1	7.00	01/16/2025	08:22
5	CCV1	1	Water	NA	NA	20.2	2.01	01/16/2025	08:25
6	Q1092-01	1	Water	NA	NA	20.1	10.15	01/16/2025	08:35
7	Q1094-03	1	Water	NA	NA	20.5	9.23	01/16/2025	08:45
8	Q1096-01	1	Water	NA	NA	20.8	7.34	01/16/2025	08:55
9	Q1096-01DUP	1	Water	NA	NA	20.9	7.35	01/16/2025	08:56
10	CCV2	1	Water	NA	NA	20.2	12.02	01/16/2025	09:00

WORKLIST(Hardcopy Internal Chain)

UB 134300

WorkList Name : ph q1096 water WorkList ID : 186926 Department : Wet-Chemistry Date : 01-16-2025 07:47:16

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1092-01	427	Water	pH	Cool 4 deg C	PSEG03	N41	01/15/2025	9040C
Q1094-03	3181	Water	pH	Cool 4 deg C	PSEG03	N41	01/15/2025	9040C
Q1096-01	D3746	Water	pH	Cool 4 deg C	PSEG03	N41	01/15/2025	9040C

Date/Time 01.16.25 08:00
 Raw Sample Received by: SA GSC
 Raw Sample Relinquished by: CL GR

Date/Time 01.16.25 11:00
 Raw Sample Received by: CL GR
 Raw Sample Relinquished by: SA GSC