


LB1343#8

7471B

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	-0.001	-0.001	67	0.000	67					4.0
0.2	0.200	0.226	0.026	1276	0.0 %	1276					11
2.5	2.500	2.416	-0.084	12925	0.0 %	12925					3
5.0	5.000	5.126	0.126	27343	0.0 %	27343					3
7.5	7.500	7.414	-0.086	39513	0.0 %	39513					1
10.0	10.000	10.019	0.019	53370	0.0 %	53370					0

LB134318 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 17 Jan 2025 08:09:54

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	67	-	0.0000	7471B	PPB	17 Jan 2025 08:27:27	S	Std
0.2 - 1	50.2	1276	-	0.2000	7471B	PPB	17 Jan 2025 08:29:44	S	Std
2.5 - 1	52.5	12925	-	2.5000	7471B	PPB	17 Jan 2025 08:32:01	S	Std
5.0 - 1	55	27343	-	5.0000	7471B	PPB	17 Jan 2025 08:34:17	S	Std
7.5 - 1	55	39513	-	7.5000	7471B	PPB	17 Jan 2025 08:39:23	S	Std
10.0 - 1	55	53370	-	10.0000	7471B	PPB	17 Jan 2025 08:47:15	S	Std
ICV59 - 1	ICV59	19502	3.6523	-	7471B	PPB	17 Jan 2025 08:50:02	U	SMPL
ICB59 - 1	ICB59	-128	-0.0377	-	7471B	PPB	17 Jan 2025 08:52:16	U	SMPL
CCV41 - 1	CCV41	27458	5.1479	-	7471B	PPB	17 Jan 2025 08:56:50	U	SMPL
CCB41 - 1	CCB41	-186	-0.0486	-	7471B	PPB	17 Jan 2025 08:59:09	U	SMPL
CRA - 1	CRA	1255	0.2222	-	7471B	PPB	17 Jan 2025 09:01:24	U	SMPL
HighStd - 1	HighStd	54129	10.1614	-	7471B	PPB	17 Jan 2025 09:03:39	U	SMPL
ChkStd - 1	ChkStd	36549	6.8568	-	7471B	PPB	17 Jan 2025 09:06:01	U	SMPL
PB166093BL - 1	PBS	-251	-0.0609	-	7471B	PPB	17 Jan 2025 09:08:23	U	SMPL
PB166093BS - 1	LCSS	23455	4.3954	-	7471B	PPB	17 Jan 2025 09:10:39	U	SMPL
Q1079-01 - 1	TAPIAL3-SB04D-R-10-010925-00-T1	542	0.0882	-	7471B	PPB	17 Jan 2025 09:15:16	U	SMPL
Q1079-01MS - 1	TAPIAL3-SB04D-R-10-010925-00-T1MS	20864	3.9083	-	7471B	PPB	17 Jan 2025 09:17:34	U	SMPL
Q1079-01MSD - 1	TAPIAL3-SB04D-R-10-010925-00-T1MSD	21178	3.9673	-	7471B	PPB	17 Jan 2025 09:19:53	U	SMPL
Q1093-01 - 1	RT-3407	6875	1.2787	-	7471B	PPB	17 Jan 2025 09:22:12	U	SMPL
CCV42 - 1	CCV42	27090	5.0787	-	7471B	PPB	17 Jan 2025 09:24:30	U	SMPL
CCB42 - 1	CCB42	-196	-0.0505	-	7471B	PPB	17 Jan 2025 09:26:49	U	SMPL
Q1094-01 - 1	72-11938	2787	0.5102	-	7471B	PPB	17 Jan 2025 09:29:04	U	SMPL
Q1098-01 - 1	OILY-STONE-COMP	856	0.1472	-	7471B	PPB	17 Jan 2025 09:31:21	U	SMPL
Q1099-01 - 1	OR-02-01152025	11867	2.2171	-	7471B	PPB	17 Jan 2025 09:33:38	U	SMPL
Q1079-01DUP - 1	TAPIAL3	455	0.0719	-	7471B	PPB	17 Jan 2025 09:36:23	U	SMPL
Q1079-01LX5 - 1		141	0.0128	-	7471B	PPB	17 Jan 2025 09:38:38	U	SMPL
Q1079-01A - 1		20831	3.9021	-	7471B	PPB	17 Jan 2025 09:40:55	U	SMPL
CCV43 - 1	CCV43	27231	5.1052	-	7471B	PPB	17 Jan 2025 09:43:10	U	SMPL
CCB43 - 1	CCB43	-134	-0.0389	-	7471B	PPB	17 Jan 2025 09:45:30	U	SMPL