

LB134324


7470A

INSTRUMENT ID: CVI

Linear

 μ Abs.:

44767



A= 0.0000e+000

B= 2.1875e-004

C= -3.6205e-002

Rho= 0.9989519

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0 D
0.0	0.000	-0.004	-0.004	148	0.000	148	0				
0.2	0.200	0.209	0.009	1119	0.0 %	1119					
2.5	2.500	2.384	-0.116	11062	0.0 %	11062					
5.0	5.000	5.044	0.044	23225	0.0 %	23225					
7.5	7.500	7.810	0.310	35870	0.0 %	35870					
10.0	10.000	9.757	-0.243	44767	0.0 %	44767					

0/0 D
 4
 -5
 1
 4
 -2

LB134324 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 17 Jan 2025 11:29:34

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	148	-	0.0000	7470A	PPB	17 Jan 2025 11:31:54	S	Std
0.2 - 1	50-2	1119	-	0.2000	7470A	PPB	17 Jan 2025 11:34:10	S	Std
2.5 - 1	52-5	11062	-	2.5000	7470A	PPB	17 Jan 2025 11:36:27	S	Std
5.0 - 1	55	23225	-	5.0000	7470A	PPB	17 Jan 2025 11:38:44	S	Std
7.5 - 1	53-5	35870	-	7.5000	7470A	PPB	17 Jan 2025 11:48:52	S	Std
10.0 - 1	51W	44767	-	10.0000	7470A	PPB	17 Jan 2025 11:51:07	S	Std
ICV60 - 1	ICV60	19691	4.2713	-	7470A	PPB	17 Jan 2025 11:58:32	U	SMPL
ICB60 - 1	ICB60	-63	-0.0500	-	7470A	PPB	17 Jan 2025 12:00:47	U	SMPL
CCV44 - 1	CCV44	23912	5.1946	-	7470A	PPB	17 Jan 2025 12:03:04	U	SMPL
CCB44 - 1	CCB44	-66	-0.0506	-	7470A	PPB	17 Jan 2025 12:05:19	U	SMPL
CRA - 1	CRA	1117	0.2081	-	7470A	PPB	17 Jan 2025 12:07:35	U	SMPL
HighStd - 1	HighStd	45603	9.9396	-	7470A	PPB	17 Jan 2025 12:09:50	U	SMPL
ChkStd - 1	ChkStd	31921	6.9466	-	7470A	PPB	17 Jan 2025 12:12:06	U	SMPL
PB166110BL - 1	PBW	92	-0.0161	-	7470A	PPB	17 Jan 2025 12:17:04	U	SMPL
PB166110BS - 1	LCSW	18477	4.0057	-	7470A	PPB	17 Jan 2025 12:19:19	U	SMPL
Q1092-01 - 1	427	1517	0.2956	-	7470A	PPB	17 Jan 2025 12:21:35	U	SMPL
Q1092-01DUP - 1	427DUP	1814	0.3606	-	7470A	PPB	17 Jan 2025 12:23:53	U	SMPL
Q1092-01MS - 1	427MS	15238	3.2972	-	7470A	PPB	17 Jan 2025 12:26:12	U	SMPL
Q1092-01MSD - 1	427MSD	14049	3.0371	-	7470A	PPB	17 Jan 2025 12:31:13	U	SMPL
Q1094-03 - 1	3181	293	0.0279	-	7470A	PPB	17 Jan 2025 12:33:28	U	SMPL
CCV45 - 1	CCV45	22474	4.8801	-	7470A	PPB	17 Jan 2025 12:35:45	U	SMPL
CCB45 - 1	CCB45	-53	-0.0478	-	7470A	PPB	17 Jan 2025 12:38:00	U	SMPL
Q1109-01 - 1	TAPIAL1-MW04I-011525-00-T3	558	0.0859	-	7470A	PPB	17 Jan 2025 12:40:17	U	SMPL
Q1109-02 - 1	TAPIAL1-MW04S-011525-00-T2	432	0.0583	-	7470A	PPB	17 Jan 2025 12:42:32	U	SMPL
Q1092-01LX5 - 1		306	0.0307	-	7470A	PPB	17 Jan 2025 12:44:48	U	SMPL
Q1092-01A - 1		17289	3.7458	-	7470A	PPB	17 Jan 2025 12:47:03	U	SMPL
CCV46 - 1	CCV46	24509	5.3252	-	7470A	PPB	17 Jan 2025 12:49:19	U	SMPL
CCB46 - 1	CCB46	-107	-0.0596	-	7470A	PPB	17 Jan 2025 12:51:36	U	SMPL