

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB134337

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/20/2025	08:30
2	CAL2	1	Water	NA	NA	20.2	7.00	01/20/2025	08:31
3	CAL3	1	Water	NA	NA	20.2	10.02	01/20/2025	08:33
4	ICV	1	Water	NA	NA	20.2	7.01	01/20/2025	08:35
5	CCV1	1	Water	NA	NA	20.2	2.01	01/20/2025	08:38
6	Q1122-01	1	Water	NA	NA	20.6	4.35	01/20/2025	08:45
7	Q1122-01DUP	1	Water	NA	NA	20.7	4.36	01/20/2025	08:46
8	CCV2	1	Water	NA	NA	20.3	12.02	01/20/2025	08:50

WORKLIST(Hardcopy Internal Chain)

JB 134337

WorkList Name : ph q1122

WorkList ID : 187001

Department : Wet-Chemistry

Date : 01-20-2025 08:21:01

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1122-01-13	RW10A-20250116	Water	pH	Cool 4 deg C	TETR06	E11	01/16/2025	9040C

Date/Time 01/20/25 08:25

Raw Sample Received by: JH WCI

Raw Sample Relinquished by: JH WCI

Date/Time 01/20/25 12:30

Raw Sample Received by: JH WCI

Raw Sample Relinquished by: JH WCI