

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB134338

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/20/2025	08:30
2	CAL2	1	Water	NA	NA	20.2	7.00	01/20/2025	08:31
3	CAL3	1	Water	NA	NA	20.2	10.02	01/20/2025	08:33
4	ICV	1	Water	NA	NA	20.2	7.01	01/20/2025	08:35
5	CCV1	1	Water	NA	NA	20.1	2.02	01/20/2025	09:00
6	Q1123-01	1	Solid	20.02	20	20.4	7.08	01/20/2025	09:10
7	Q1123-01DUP	1	Solid	20.03	20	20.5	7.09	01/20/2025	09:11
8	CCV2	1	Water	NA	NA	20.1	12.02	01/20/2025	09:15

WORKLIST(Hardcopy Internal Chain)

WP 134328

WorkList Name : ph q1123

WorkList ID : 187000

Department : Wet-Chemistry

Date : 01-20-2025 08:19:47

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1123-01 ~ B	BU-03-1172025	Solid	pH	Cool 4 deg C	PSEG05	E11	01/17/2025	9045D

Date/Time 01/20/25 08:25

Raw Sample Received by: PP WCI

Raw Sample Relinquished by: CP SM

Date/Time 01/20/25

Raw Sample Received by: CP SM

Raw Sample Relinquished by: BLADE