

LB134354


245.1

INSTRUMENT ID: CV2

Linear

 μ Abs.:

33356



A= 0.0000e+000

B= 2.9649e-004

C= -6.3418e-002

Rho= 0.9995976

Accept = Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	6/10
0.0	0.000	-0.039	-0.039	81	0.000	81					
0.05	0.050	0.043	-0.007	360	0.0 %	360	0				-14
0.2	0.200	0.197	-0.003	880	0.0 %	880					-2
2.5	2.500	2.494	-0.006	8626	0.0 %	8626					0
5.0	5.000	5.015	0.015	17127	0.0 %	17127					0
7.5	7.500	7.714	0.214	26230	0.0 %	26230					3
10.0	10.000	9.826	-0.174	33356	0.0 %	33356					-2

LB134354 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 21 Jan 2025 10:54:02

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	81	-	0.0000	245.1	PPB	21 Jan 2025 14:55:26	S	Std
0.05 - 1	50.45	360	-	0.0500	245.1	PPB	21 Jan 2025 14:57:43	S	Std
0.2 - 1	50.2	880	-	0.2000	245.1	PPB	21 Jan 2025 15:00:01	S	Std
2.5 - 1	52.5	8626	-	2.5000	245.1	PPB	21 Jan 2025 15:02:18	S	Std
5.0 - 1	55	17127	-	5.0000	245.1	PPB	21 Jan 2025 15:04:35	S	Std
7.5 - 1	57.5	26230	-	7.5000	245.1	PPB	21 Jan 2025 15:06:51	S	Std
10.0 - 1	57.1	33356	-	10.0000	245.1	PPB	21 Jan 2025 15:18:01	S	Std
ICV64 - 1	ICV64	13496	3.9386	-245.1		PPB	21 Jan 2025 15:26:17	U	SMPL
ICB64 - 1	ICB64	-98	-0.0925	-245.1		PPB	21 Jan 2025 15:28:32	U	SMPL
CCV57 - 1	CCV57	18478	5.4151	-245.1		PPB	21 Jan 2025 15:30:49	U	SMPL
CCB57 - 1	CCB57	-67	-0.0833	-245.1		PPB	21 Jan 2025 15:33:04	U	SMPL
CRA - 1	CRA	373	0.0472	-245.1		PPB	21 Jan 2025 15:35:21	U	SMPL
HighStd - 1	HighStd	33388	9.8358	-245.1		PPB	21 Jan 2025 15:37:37	U	SMPL
ChkStd - 1	ChkStd	24318	7.1467	-245.1		PPB	21 Jan 2025 15:39:52	U	SMPL
PB166158BL - 1	PBW	60	-0.0456	-245.1		PPB	21 Jan 2025 15:44:51	U	SMPL
PB166158BS - 1	LCSW	13282	3.8746	-245.1		PPB	21 Jan 2025 15:47:06	U	SMPL
P3390-10 - 1	LOD-MDL-WATER	374	0.0475	-245.1		PPB	21 Jan 2025 15:58:00	U	SMPL
P3390-10RE - 1	LOD-MDL-WATER-RE	778	0.1673	-245.1		PPB	21 Jan 2025 16:00:16	U	SMPL
P3390-12 - 1	MDL-WATER-06-QT3-2024	389	0.0519	-245.1		PPB	21 Jan 2025 16:02:33	U	SMPL
CCV58 - 1	CCV58	18250	5.3475	-245.1		PPB	21 Jan 2025 16:04:49	U	SMPL
CCB58 - 1	CCB58	124	-0.0267	-245.1		PPB	21 Jan 2025 16:09:34	U	SMPL