

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Iwona

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB134382

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3157

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	01/23/2025	12:10
2	Q1140-01		1	13	100.00	>212.0	>212.0	01/23/2025	12:40
3	Q1140-01DUP		1	14	100.00	>212.0	>212.0	01/23/2025	13:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 134382

WorkList Name : FLASH POINT-012325

WorkList ID : 187088

Department : Wet-Chemistry

Date : 01-23-2025 09:22:44

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1140-01	FRAC-TANK-F06078	Water	Flash Point	Cool 4 deg C	PSEG03	E11	01/21/2025	1010B

Date/Time 01/23/25 11:30
Raw Sample Received by: 12 (WC)
Raw Sample Relinquished by: [Signature] 12 (WC)

Date/Time 01/23/25 13:20
Raw Sample Received by: [Signature] 12 (WC)
Raw Sample Relinquished by: [Signature] 12 (WC)