

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB134387

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/23/2025	14:25
2	CAL2	1	Water	NA	NA	20.2	7.00	01/23/2025	14:26
3	CAL3	1	Water	NA	NA	20.3	10.02	01/23/2025	14:30
4	ICV	1	Water	NA	NA	20.3	7.01	01/23/2025	14:35
5	CCV1	1	Water	NA	NA	20.3	2.01	01/23/2025	14:40
6	Q1140-01	1	Water	NA	NA	20.6	8.26	01/23/2025	14:49
7	Q1140-01DUP	1	Water	NA	NA	20.7	8.27	01/23/2025	14:50
8	Q1165-01	1	Water	NA	NA	20.6	7.76	01/23/2025	15:00
9	Q1167-02	1	Water	NA	NA	20.3	7.44	01/23/2025	15:05
10	CCV2	1	Water	NA	NA	20.2	12.02	01/23/2025	15:10

WORKLIST(Hardcopy Internal Chain)

134387

WorkList Name : ph w q1167 WorkList ID : 187106 Department : Wet-Chemistry Date : 01-23-2025 14:02:50

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1140-01	FRAC-TANK-F06078	Water	pH	Cool 4 deg C	PSEG03	E11	01/21/2025	9040C
Q1165-01	91624-F	Water	pH	Cool 4 deg C	PSEG03	N41	01/23/2025	9040C
Q1167-02	34784	Water	pH	Cool 4 deg C	PSEG03	N41	01/23/2025	9040C

Date/Time 01/23/25 14:15
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 01/23/25 16:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]