

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB134393

Slope : 98.4

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/24/2025	08:10
2	CAL2	1	Water	NA	NA	20.1	7.00	01/24/2025	08:11
3	CAL3	1	Water	NA	NA	20.2	10.02	01/24/2025	08:12
4	ICV	1	Water	NA	NA	20.3	7.01	01/24/2025	08:15
5	CCV1	1	Water	NA	NA	20.2	2.01	01/24/2025	08:17
6	Q1167-01	1	Solid	20.02	20	20.6	6.63	01/24/2025	08:25
7	Q1167-01DUP	1	Solid	20.03	20	20.7	6.64	01/24/2025	08:26
8	CCV2	1	Water	NA	NA	20.3	12.02	01/24/2025	08:30

WORKLIST(Hardcopy Internal Chain)

WP 134393

WorkList Name : ph q1167 s

WorkList ID : 187110

Department : Wet-Chemistry

Date : 01-24-2025 07:58:37

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1167-01	34785	Solid	pH	Cool 4 deg C	PSEG03	N41	01/23/2025	9045D

Date/Time 01/24/25 08:00
 Raw Sample Received by: JB COC
 Raw Sample Relinquished by: JB COC

Date/Time 01/24/25 10:00
 Raw Sample Received by: JB COC
 Raw Sample Relinquished by: JB COC