

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB134408

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/24/2025	13:00
2	CAL2	1	Water	NA	NA	20.2	7.00	01/24/2025	13:01
3	CAL3	1	Water	NA	NA	20.3	10.02	01/24/2025	13:05
4	ICV	1	Water	NA	NA	20.3	7.01	01/24/2025	13:10
5	CCV1	1	Water	NA	NA	20.2	2.01	01/24/2025	13:15
6	Q1169-02	1	Solid	20.02	20	20.1	7.74	01/24/2025	13:25
7	Q1169-02DUP	1	Solid	20.03	20	20.2	7.75	01/24/2025	13:26
8	Q1169-04	1	Solid	20.02	20	20.1	9.20	01/24/2025	13:37
9	Q1169-06	1	Solid	20.03	20	20.2	7.64	01/24/2025	13:45
10	CCV2	1	Water	NA	NA	20.2	12.02	01/24/2025	13:50

WORKLIST(Hardcopy Internal Chain)

134408

WorkList Name : corrosivity q1162

WorkList ID : 187137

Department : Wet-Chemistry

Date : 01-24-2025 11:24:25

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1169-02	ARS20-0006	Solid	Corrosivity	Cool 4 deg C	PSEG03	N51	01/23/2025	9045D
Q1169-04	REGULATOR-BUILDING-GRAV	Solid	Corrosivity	Cool 4 deg C	PSEG03	N51	01/23/2025	9045D
Q1169-06	SPIILL-SITE	Solid	Corrosivity	Cool 4 deg C	PSEG03	N51	01/23/2025	9045D

Date/Time 01/24/25 12:30

Raw Sample Received by: SO WOC

Raw Sample Relinquished by: JTC

Date/Time 01/24/25 15:30

Raw Sample Received by: JTC

Raw Sample Relinquished by: SO WOC